

# Key technical aspects and vascular access safety in membrane-based therapeutic plasma exchange for the pediatric population

---

## Keywords

children, filter, central catheter, Therapeutic plasma exchange, technical adverse effects, clinical adverse effects

---

## Abstract

### Introduction

Therapeutic plasma exchange (TPE) is a widely used extracorporeal blood purification procedure. While its principles are similar across age groups, the technical complexity and risks are heightened in pediatric patients due to unique technical considerations. This study assesses the safety regarding both technical and clinical complications of vascular access and filters of membrane technique TPE performed in a pediatric population at a single center.

### Material and methods

This retrospective cohort study reviewed charts of patients undergoing TPE at a level 3 referral center over 25 years.

### Results

The study involved 178 patients undergoing a total of 740 procedures during 214 sessions, predominantly using femoral vascular access. The median duration of the entire TPE sessions (treated as a surrogate for catheter lifespan) was 118.5 hours. Technical complications occurred in 20.8% of procedures (31.6% of events), while clinical complications occurred in 4.2% (4.7% of events). Technical complications were nearly five times more frequent than clinical complications, with technical events being 6.7 times more common. The proportion of patients experiencing clinical complications amounted to 15.2%. Logistic regression demonstrated that each additional day of catheter placement increased the probability of technical complications by 3%. Additionally, each year of patient age decreased the likelihood of clinical complications by 8.9%, while Fresh Frozen Plasma (FFP) supplementation within a session increased the probability of clinical complications by 21.6%.

### Conclusions

Prolonged catheter use increases technical events, while FFP supplementation elevates catheter-related clinical complication risk. Advancing patient age reduces the likelihood of clinical complications, underscoring age-specific safety considerations in pediatric TPE.

## **Introduction**

Therapeutic plasma exchange (TPE) is an extracorporeal blood purification procedure employed in clinical settings. It functions on the principle of eliminating dissolved pathogenic entities, such as autoantibodies or immune complexes, from the blood plasma while simultaneously replenishing essential plasma components. This technique is applied across a broad spectrum of medical disciplines, particularly in neurology, hematology, rheumatology, metabolic disorders, nephrology, and selected aspects of toxicology [1].

The fundamental principles of TPE are consistent between adults and children; however, several technical distinctions exist, including the establishment of vascular access, differences in volume of distribution, and increased vascular complications associated with smaller blood vessels. Additionally, the lack of patient cooperation during the procedure contributes to the overall risk and technical complexity, making this form of therapy more challenging in children than in adults [2]. Although there are studies on the clinical indications and complications associated with this technique in the pediatric population, research focusing solely on technical issues in this field remains somewhat scarce. Hence, we decided to analyze our 25-year experience with TPE in children, with a particular emphasis on vascular access, filter issues, and anticoagulation challenges in this heterogeneous patient population.

## **Aim of the study**

The objective of this study was to assess the safety of vascular access – central venous catheters (CVC), filters, anticoagulation, and the performance of membrane technique components in TPE conducted over a 25-year period in a pediatric population with neurological and non-neurological conditions at a single tertiary referral hospital.

## **Material and Methods**

This study is a retrospective chart review of patients who were qualified for the TPE procedure due to neurological and non-neurological conditions, performed at the Pediatric Nephrology and

Hypertension Clinic in *(blinded for review)* from January 1998 to December 2022. The non-neurological group included patients from the intensive care, pediatric nephrology, and pediatric hematology departments.

The local ethics committee approved the study (consent reference number 118.6120.187.2023) and informed consent was waived due to its retrospective nature. The study conducted a detailed analysis of data from the medical histories and TPE charts of each patient included in the study. A TPE session was defined as a series of TPE procedures performed on a patient with less than a four-week interval between each procedure. If the interval between TPE procedures was four or more weeks, then the TPE sessions were considered separate. This is related to the maximum duration of use for an acute dual-lumen catheter.

Due to the inability to obtain precise data regarding the time of implantation of the dual-lumen CVC for TPE, for statistical purposes in this review, the duration of the session has been adopted as a surrogate for the lifespan of the given catheter. If a single TPE procedure was conducted within a given session, the duration of that procedure equated to the CVC's utilization time.

For statistical calculations, anthropometric data (age, body mass, height and corresponding percentile values) recorded at the beginning of the given TPE session were considered. The categorization of indication for performing TPE for each patient was determined according to the most recent guidelines of The American Society for Apheresis (ASFA) [3].

The decision to qualify a patient for treatment using the TPE method was made based on the primary diagnosis by the specialist neurologist, nephrologist, intensivist, or pediatric hematologist, while the decision regarding the placement of an acute dual-lumen catheter and its size was made after consultation between the nephrologist and an experienced surgeon. Dual-lumen catheters ranging from 8F to 12.5F were individually adapted to the child's morphology and body weight according to literature recommendations [4,5].

TPE procedures were carried out using the filtration method in accordance with the prevailing guidelines, utilizing the following types of machines and filters: 1) Hospal machine with filters: PSN1, Hemaflex BT 900, PF1000N (years: 1998 - 2008); 2) Prisma machine with filters: PF1000N, PF2000N (years: 2009-2015); and 3) Prisma Flex machine with filters: PF1000N (from 2016 onwards). The sizes of the specific filters were chosen according to the child's body mass in line with the medical product's characteristics.

The TPE procedures were performed according to our center's protocol, which included, among others: the establishment of appropriate vascular access for age (i.e., catheter size and location), priming and anticoagulation of the circuit, laboratory tests, and continuous monitoring of the patients (including hemoglobin oxygen saturation, heart rate, and blood pressure values) during the TPE procedure [5,6].

As supplements, fresh frozen plasma (FFP), 5% human albumin (HA) solution, 6% hydroxyethyl starch (HES) solution, and crystalloids were used. The type and configuration of supplements used depended on both the clinical condition of the patient and the timeline in which the treatment was administered [5,6,7]. The nephrology specialist planned and supervised the course of the TPE procedures in coordination with the team of specialists managing the patient's care.

The initial total plasma volume (TPV) to be exchanged (Estimated Plasma Volume status - EPV) was calculated based on the patient's hematocrit (Hct) level and body weight using the Kaplan formula:  $EPV = [0.065 \times \text{body weight (kg)}] \times [1 - \text{Hct}]$  [3,5]. Unfractionated heparin was used for anticoagulation, with an initial dose of 50-70 IU/kg body weight followed by a continuous intravenous infusion (and/or boluses) of unfractionated heparin at a dose of 10-30 IU/kg/h, with dosing adjusted to achieve a therapeutic activated clotting time (ACT) between 180 - 240 seconds [6]. The analysis included, among other things: the number of performed plasmapheresis procedures and technical aspects of the TPE process such as: blood flow rate, duration of each procedure, size of the dual-lumen catheter, type of filter, initial dose of anticoagulant, ACT times values, as well as clinical and

technical complications resulting from the used vascular access and membrane technique, and the actions taken in response to the observed complications.

## **Statistical Analysis**

Statistical analysis was performed utilizing MATLAB software (The MathWorks Inc., 2022b; MATLAB version 9.13.0 (R2022b), Natick, Massachusetts, USA). Data were articulated as means  $\pm$  standard deviations (SD) or medians with interquartile ranges (IQR). Distribution normality was assessed using the Shapiro-Wilk test. Dependent upon the distribution of the variables, the following tests were employed: Student's t-test, rank-sum Wilcoxon test, Fisher's exact test or McNemar's test for matched samples, and Pearson's linear correlation. For multivariate analysis, a generalized multivariate linear model (GLM) with backward elimination was utilized. The receiver operating characteristic (ROC) analysis was also implemented providing the relevant parameters. A p-value of less than 0.05 was pre-determined as the threshold for statistical significance.

## **Results**

The study cohort contained of 178 patients who collectively underwent 740 procedures during 214 sessions.

The neuroimmunological population (NE) consisted of 4 subgroups: 1) Acute inflammatory demyelinating polyradiculoneuropathy (AIDP; Guillain- Barré syndrome, GBS): 65 children, 65 sessions, 247 TPE procedures, averaging 3.8 procedures per patient; 2) Polyneuropathy (PN): 5 children, 5 sessions, 19 TPE procedures, averaging 3.8 procedures per patient; 3) Myasthenia gravis (MG): 8 patients, 13 sessions, 56 TPE procedures, averaging 1.63 sessions per patient and 7 procedures per patient; 4) Multiple sclerosis (MS): 3 children, 24 sessions, 38 TPE procedures, averaging 8 sessions per patient and 12.67 procedures per patient. In total, the neuroimmunological group included 81 patients; 360 therapeutic plasma exchanges were performed during 107 sessions. The non-neuroimmunological population (non-NE) consisted of patients from the 7 main diagnostic fields: 1) pediatric intensive care unit (PICU) patients: 13 children, 13 sessions, 30 TPE procedures,

averaging 2.3 procedures per patient; 2) toxicological patients: 21 children, 21 sessions, 39 TPE procedures, 3.3 procedures per patient; 3) hematology: 12 children, 13 sessions, 42 TPE procedures, 3.5 procedures per patient; 4) nephrology – Rapid progressive glomerulonephritis: 11 children, 13 sessions, 71 TPE procedures, 6.5 procedures per patient; 5) nephrology - Systemic lupus erythematosus: 13 children, 15 sessions, 70 TPE procedures, 5.4 procedures per patient; 6) nephrology - Focal segmental glomerulosclerosis: 4 children, 6 sessions, 28 TPE procedures, 7 procedures per patient; 7) nephrology - Thrombotic microangiopathy: 23 children, 26 sessions, 100 TPE procedures, 4.4 procedures per patient. In total, the non-neuroimmunological group included 97 patients; 380 therapeutic plasma exchanges were performed during 107 sessions.

Detailed epidemiological data on the studied population of NE and non-NE patients, as well as the composition and doses of used supplements, are included in Tables 1 and Supplementary Table 1.

Regarding epidemiological data – only body mass in the NE group was significantly higher. In the non-NE group, procedures using FFP were performed significantly more often (88.2% vs 76.1%), but the dose was significantly smaller (30 vs 33.6 ml/kg) than in the NE group. Similarly, the duration of TPE procedures and the rate of supplement exchanges did not differ.

In both patient groups, the femoral vascular access was predominant (in the NE group 86.9% vs 60.7% in the non-NE group). Due to potential complications (including pneumothorax, vascular narrowing) – at our center, the use of short-term subclavian vein access is avoided. CVCs were mostly inserted in the femoral vein, which is the preferred site of insertion in acute hemodialysis/TPE due to a smaller number of complications [8]. Detailed data regarding vascular access, its location, and size are included in Supplementary Table 2.

Dysfunction of the catheter was defined as a failure to attain sufficient extracorporeal blood flow for an efficient procedure. An exit site infection (ESI) was defined as signs of inflammation in the area surrounding the catheter exit site and/or the presence of exudate that proves to be culture-positive.

Detailed information regarding catheter-related technical and clinical complications, interventions

conducted in response to them, and issues related to the filter, machine, and reasons for premature termination of the TPE procedures are included in Tables 2, 3, and 4.

Comparing the studied populations - in the NE group, a statistically significant 2.3-fold higher incidence of TPE with technical complications (29.4% vs 12.6%) and a 2.4-fold higher rate of technical complication events (45.3% vs 18.7%) were noted (including a 2-fold higher rate of procedures on reversed lines and a 3.4-fold greater incidence of malfunctions in the arterial part of the catheter) compared to the non-NE population.

However, the incidence of TPE with clinical complications (3.9% vs 4.5%), and the percentage of patients with clinical complications (13.6% vs 16.5%) were comparable between the groups.

Interestingly, it was found that in the NE group there was a statistically significant 1.7-fold higher percentage of patients requiring intervention due to technical and clinical complications related to vascular access (29.6% vs 17.5%); 1.8-fold higher rate of TPE with technical complications related to the filter (6.7% vs 3.7%); 1.8-fold higher rate of adverse events (AEs) related to the filter (11.7% vs 6.6%); 2.75-fold higher rate of TPE with premature termination (4.4% vs 1.6%) and a 3.8-fold higher rate of AEs related to premature termination of TPE (6.1% vs 1.6%) compared to the non-NE group.

In the analysis of the study population concerning the administration of FFP during TPE, it was found that the group receiving FFP (FFP1) exhibited a significantly higher incidence of TPE with technical events (23% vs 10.7%; a 2.1-fold increase) and clinical events (4.9% vs 0.8%; a 6.1-fold increase).

Additionally, the occurrence of clinical complications was 7- times higher in the FFP1 group compared to those not receiving FFP (FFP0, 5.6% vs 0.8%).

In the FFP1 group, a statistically significant 2.4-fold increase in the incidence of TPE with any event (clinical and technical) related to vascular access was also observed (27.9%) compared to the FFP0 group (11.5%;  $p < 0.05$ ; one-sided Fisher's test). The analyzed data therefore indicate that the use of FFP contributes to a higher percentage of clinical AEs in this group of patients.

Detailed information on catheter-related technical and clinical complications, challenges related to

the filter, machine, and causes of premature termination of the TPE procedure in the FFP1 and FFP0 subgroups are included in Tables 5 and 6.

In the entire studied population, the incidence of TPE with technical complications was 20.8% (154/740), and the rate of all technical complication events was 31.6% (234/740); this equates to 0.9 TPEs with a technical complication and 1.3 technical complication events per patient. Meanwhile, the incidence of TPE with clinical complications was 4.2% (31/740), and the rate of all clinical complication events was 4.7% (35/740); the percentage of patients with clinical complications was 15.2% (27/178) and the rate of all clinical complication events in the studied population was 19.7% (35/178). Therefore, in the studied population, the frequency of TPE with technical complications is nearly 5 times greater than with clinical complications (20.8% vs 4.2%;  $p < 0.05$ ); similarly, the frequency of technical events is 6.7 times greater than that of clinical events (31.6% vs 4.7%;  $p < 0.05$ ).

The median duration of entire TPE sessions (which in this study serves as a surrogate for catheter lifespan) was nearly three times longer in the NE group compared to the non-NE group (145 vs 49 hours). Across the entire studied population, it averaged 118.5 hours. It's important to underline, that in cases where a patient was diagnosed with severe kidney injury or end-stage kidney disease (ESKD) (7 nephrological patients, in 7 sessions) or when longer period of TPE was planned in advance (2 neurological patients in 2 sessions) - a permanent catheter was electively used in such cases.

Nevertheless, narrowing the studied population to acute catheters only, their usage time was also 3.1 times longer in the NE population compared to the non-NE population, with the median usage time for these catheters in the entire population being 99 hours.

Furthermore, the incidence of TPE with any clinical and technical event related to vascular access was significantly higher, by 1.9 times in the NE group (33.3%) compared to the non-NE group (17.1%). Across the whole study population, this rate was 25%. Thus, the significantly longer duration of catheter use in the NE group may influence the higher number of technical AEs and the proportion of events related to the filter in this group of patients.



Furthermore, a multivariate logistic regression analysis of the entire sessions on the occurrence of catheter-related technical complications in a given TPE session showed that belonging to the NE group increases the probability of a technical complication by 2.4 times (2.1 for acute catheters only), and each day the catheter is in place increases the probability of a technical complication by 3% (6% for acute catheters only). Moreover, a multivariate logistic regression analysis of the entire sessions on the occurrence of catheter-related clinical complications in a given TPE session indicated that each additional year of the patient's life decreases the probability of a clinical complication by 8.9% (9.1% for acute catheters only), and each additional FFP procedure during the session increases the probability of a clinical complication by 21.6% (20.8% for acute catheters only) (Table 7). Additionally, an analysis of AUROCs conducted for entire TPE sessions (summing the number of AEs occurring in each session) revealed that the AUROC for the selected cut-off value of 5 days of catheter lifespan for the occurrence of any catheter-related technical AEs was 71% (likelihood ratio (LR): 2.16; sensitivity: 76.5%; specificity: 64.7%; CI 95%: 65.26-76.65; p-value <0.05), indicating good predictive value. This model suggests a 71% probability that the model will correctly distinguish between a technical AEs and no event. Similarly, an analysis of AUROCs conducted for entire TPE sessions (summing the number of AEs occurring in each session) showed that the AUROC for the selected cut-off value of 5 days of catheter lifespan for the occurrence of TPE with any vascular access-related technical event was 71.3% (LR: 2.15; sensitivity: 79.9%; specificity: 62.9%; CI 95%: 65.29-77.34; p-value <0.05), also indicating good predictive value. This model suggests a 71.3% probability that the model will correctly distinguish between a technical AEs and no event. The AUROC values for the above data, extracted only for the subpopulation of acute catheters, were below 70%; therefore, they were not included in this paper.

## Discussion

Therapeutic plasma exchange remains a widely used treatment modality for various diseases in children by removing plasma containing pathogenic agents. While the principles of TPE are similar in

adults and children, there are several technical differences that may affect the overall efficacy and safety of this treatment in the pediatric population.

One of these factors is CVC, which in the pediatric population is usually performed by interventional radiologist under direct visualization and often under general anesthesia.

In this study, technical AEs related to CVC were analyzed, including: the necessity of conducting the procedure using reversed lines (which affects the procedure's efficacy), incidents involving difficulties in blood withdrawal and return (indicating malfunction of the arterial and venous parts of the catheter, respectively), thrombosis within the catheter, spontaneous catheter repositioning, and catheter leakage.

AEs of a clinical nature related to the CVC were also scrutinized, including: thrombosis in the vessel with the catheter in place, edema of the extremity with the catheter, catheter ESI, isolated pain at the catheter implantation site, and bleeding at the implantation site of the CVC. Additionally, AEs related to the filters themselves were analyzed, such as episodes of clotting, ruptures, the necessity for replacement, and increased pressure on the filter without clotting. The severity of these AEs was indirectly assessed by analyzing the type and number of interventions required for the aforementioned AEs. Furthermore, the frequency and causes of premature termination of the TPE procedure were evaluated.

In this population, due to the specific characteristics of the pediatric population, the femoral access was the most commonly selected access site (73.8%). In the literature, femoral catheter placement is reported at levels ranging from 5.9% to 80% [8-14].

In this study, catheter-related technical malfunctions were recorded in 230 out of 740 TPE sessions (31.1%), with the two most frequently reported catheter-related technical complications being arterial catheter malfunction (15.8%) and the necessity to conduct the procedure using reversed lines (13.9%). Other noted complications were very rare (0.1–1.4% per TPE).

224 Additionally, after excluding AE related to procedures performed on reversed lines, the rate of  
225 technical events was reduced to 17.7% (131/740) per TPE. In the literature, reports on access  
226 malfunction vary widely, ranging from 1.2% to 38.8% per TPE [13,15–21]. This discrepancy may result  
227 from differences in the definition of catheter malfunction used in different centers and statistical  
228 calculation methods.

229 In this study, each type of AE was recorded separately; thus, during a single TPE session, both an  
230 arterial catheter malfunction and a procedure conducted on reversed lines could be documented.

231 Catheter leakage is reported in the literature at a level of 2% per TPE [20] and 5% per patient [22],  
232 while catheter thrombosis is reported at 1.6% per TPE [15] and at levels ranging from 10% [14] to  
233 17.4% [10] per patient. In the studied population, catheter leakage was observed in 1 case (0.1% per  
234 TPE), while catheter thrombosis occurred in only 2 patients (1.1%) and in 0.3% per procedure.

235 The two most frequently documented clinical complications were thrombosis in the catheterized  
236 vessel (1.5%) and bleeding at the implantation site of a double-lumen catheter (1.4%). Other  
237 recorded clinical complications were even rarer (0.3–0.8% per TPE). In the literature, catheter-related  
238 thrombosis is noted at levels of 0.08% [1] to 0.4% [16] per TPE and at levels of 1.7–6.25% per patient  
239 [8,16,22,23].

240 In contrast, bleeding at the implantation site of a double-lumen catheter aligns with data from the  
241 literature, which describes it at levels of 0.25% to 3% per TPE [1,8,17,24].

242 Pain at the catheter implantation site was observed in 6 cases (0.8% per TPE) in this study, whereas  
243 the literature notes ‘abnormal sensation’ in 0.25% per TPE [1].

244 In the literature, the rate of catheter-related infections (bacteremia/catheter-related sepsis) is  
245 documented at 4.8–17% of patients [8,16,25–27] and at 0.25–2.1% per procedure [1,16,26]. In this  
246 study, no AEs in the form of bacteremia or catheter-related sepsis were recorded. This could be  
247 attributed to the fact that children at risk of systemic infection (e.g., PICU patients, nephrological, or

hematological patients) were treated with systemic antimicrobial therapy due to their underlying disease, which a priori complicates the analysis of such data.

In this study, complications such as pneumothorax (reported in the literature at 0.9% of patients [25]) and premature catheter disconnection (reported in the literature at 1.8% per TPE [20]) were not observed. However, one case of spontaneous catheter repositioning was documented (0.1% per TPE).

The two most common medical interventions related to catheter-associated technical and clinical complications in the studied population were the systemic administration of low-molecular-weight heparin (1.8%) and the application of a surgical compression dressing (1.2%). The rate of other interventions did not exceed 1% per TPE.

The two most frequently documented technical complications related to the filter were clotting within the filter (3.9%) (reported in the literature at 19.6% per TPE [25]) and the need for filter replacement (3.2%).

The rate of prematurely terminated TPE sessions due to various causes in this study was 3% (filter-related issues: 2.4%; vascular access-related technical issues: 1%). Data reported in the literature range from 4.4 to 7% [22,28].

Machine-related problems over a 25-year period were recorded at a rate of 0.3% per TPE, whereas the literature reports this complication at a level of 18.6% per TPE [29].

The median duration of entire TPE sessions (which, in this study, serves as a surrogate for catheter lifespan) in the studied population was 118.5 hours (equivalent to 4.9 days), with a range from 0.75 hours to 1732 hours (including permanent catheters), and 100 hours (4.2 days) with the range to 767 hours (32 days) - excluding them. In the literature, the mean catheter lifespan is reported as ranging from 1 to 27 days, with an average of  $8.1 \pm 6.4$  days [8].

Analysis of the studied population revealed that a longer duration of catheter lifespan (in the NE group) may influence a higher number of catheter-related technical AEs and an increased percentage of filter-related AEs, as each additional day of catheter lifespan increases the odds of a technical

complication by 3%. Furthermore, it was observed that after five days of catheter lifespan, the probability of experiencing any catheter-related technical AE or a TPE session with any technical event rises to 71%.

The use of FFP was found to contribute to a higher percentage of catheter-related clinical AEs, as each additional FFP procedure during a session increases the odds of a clinical complication by 21.6%. Additionally, it was observed that each additional year of the patient's age reduces the odds of a clinical complication by 8.9%, which is consistent with reports in the literature stating that catheter-related problems are significantly associated with younger age [15].

Comparing technical complications of TPE in the literature is challenging due to differences in analysis methods depending on the author. Percentages are calculated relative to the total number of TPEs, the number of patients, or the total number of recorded complications. Therefore, in this study, the focus was placed on comparing selected groups of complications, presenting them in a systematic manner as the number of TPEs with a given complication and as the number of AEs relative to the total number of TPEs. For selected clinical complications, they were also presented relative to the number of patients.

Additionally, the study examined the influence of factors such as patient age, body weight, duration of the treatment session (used as a surrogate for CVC lifespan), the use (or non-use) of FFP during TPE, and group belonging (NE or Non-NE) on the analyzed complications.

For the evaluation of procedural and patient safety, we primarily recorded technical and clinical access-related and filter-related problems. In the studied population, the percentage of patients with clinical complications was 15.2%, which is consistent with data reported in the literature [26]. All of these AEs were mild in severity and could be managed with standard methods. It is also noteworthy that no severe AEs related to the TPE procedure or CVC was recorded that resulted in patient death.

The obtained data indicate a very good safety profile for the TPE procedure at our center.

The membrane-based technique might also play a role in the incidence of technical complications during TPE. Webb et al. compared membrane TPE (mTPE) and centrifugal TPE (cTPE) in 105 patients under 21 years of age and reported higher machine-related complications (17.4%) in mTPE compared to 7.1% in cTPE, as well as higher rates of circuit clotting (6.7%) vs. none in cTPE. Although they found no significant differences in patient complications between the techniques [30]. Similar findings regarding lower clotting rates in cTPE were described by Kielstein et al [31].

An important limitation of the obtained results is the fact that this study is retrospective and single center, spanning 25 years, during which the technical conditions of the procedures evolved, including changes in machines, filters, supplements, and CVCs. However, complications during TPE sessions are typically well-documented in our center, as they often require procedural adjustments or additional medication. We are also aware of the bias affecting the results of catheter-related complications, stemming from the fact that 4.2% of sessions underwent TPE procedures on a permanent catheter. Nevertheless, as we did not want to exclude these TPEs from the database due to other valuable data subjected to analysis, we assumed that these proportions were relatively small in the whole dataset that they would not have a significant impact on the overall results. Furthermore, in the multivariate logistic regression analyses – data for all permanent and separately for acute catheters was calculated.

Nonetheless, to the best of our knowledge, this study represents the most extensive patient database from a single tertiary referral center in Central and Eastern Europe, spanning such a prolonged period, and meticulously evaluating the technical safety of the TPE membrane procedures in children. As Meyer and Wong stated, there are striking differences in the rates of complications between published research on the safety of TPE in children. They also suggested that the inclusion criteria and analytical approaches might significantly differ, emphasizing the need for prospective, collaborative clinical trials, which could portray a true incidence of AE in pediatric TPE procedures [32].

The obtained data provide a basis for concluding that in the membrane TPE technique, technical complications predominate, while clinical complications are relatively rare and mild, and in our study, complications were manageable with standard interventions, with their likelihood decreasing with the patient's age.

Importantly, the data highlight key factors that can further enhance safety: awareness of potential complications allows for better preparation and response; careful selection of candidates—particularly those with known risk factors—may help avoid unnecessary risk by incorporating alternative therapies. Additionally, a longer duration of catheter lifespan is associated with a higher number of catheter-related technical AEs and an increased percentage of filter-related AEs, whereas the use of FFP contributes to a higher percentage of catheter-related clinical AEs.

## References

- [1] Lu J, Zhang L, Xia C, Tao Y. Complications of therapeutic plasma exchange: A retrospective study of 1201 procedures in 435 children. *Medicine (Baltimore)*. 2019;98(50): e18308.
- [2] Goldstein SL. Therapeutic apheresis in children: special considerations. *Semin Dial*. 2012;25(2): 165-170.
- [3] Connelly-Smith L, Alquist CR, Aquil NA, et al. Guidelines on the Use of Therapeutic Apheresis in Clinical Practice - Evidence-Based Approach from the Writing Committee of the American Society for Apheresis: The Ninth Special Issue. *J Clin Apher*. 2023;38(2): 77-278.
- [4] Zachwieja K. Ciągłe Metody Leczenia Nerkozastępczego (continuous renal replacement therapy, CRRT). In: Pietrzyk J.J, Kwinta P, editors. *Pediatrics*, Wydawnictwo UJ; 2018. vol. 2, p. 135-139.
- [5] Ismail N., Kiproff D.D., Hakim R. M. In: Daugirdas J.T., Blake P.G., Ing T.S, editors. *Podręcznik dializoterapii*, 2<sup>nd</sup> edn. Czelej; 2008. p. 176-190.
- [6] Zachwieja K. Plazmafareza. In: Pietrzyk J.J, Kwinta P., editors. *Pediatrics*, Wydawnictwo UJ; 2018. vol. 2, p. 140-142.


- [7] Levy J. Plasma Exchange. In: Johson J.R, Feehallt J., Floege J., editors. Comprehensive Clinical Nephrology, 5th edn. Elsevier Saunders; 2010. p. 1122-1130.
- [8] Rus RR, Premru V, Novljan G, Grošelj-Grenc M, Ponikvar R. Fate of Central Venous Catheters Used for Acute Extracorporeal Treatment in Critically Ill Pediatric Patients: A Single Center Experience. *Ther Apher Dial.* 2016;20(3): 308-311.
- [9] Öztürk AG, Küçük ZE, Özcan S, et al. Use of Therapeutic Plasma Exchange in the Pediatric Intensive Care Unit. *Turk Arch Pediatr.* 2022;57(2): 186-192.
- [10] Oto A, Kilic N, Kazanci EG, Akaci O, Ekici A. Therapeutic Plasma Exchange in Critically ill Pediatric Patients. *International Journal of Medical Science and Clinical Invention.* 2022;9(11): 6333–6342.
- [11] Özsoylu S, Dursun A, Çelik B. Therapeutic Plasma Exchange in Pediatric Intensive Care Unit: A Single-center Experience. *Indian J Crit Care Med.* 2021;25(10): 1189-1192.
- [12] Aygün F, Varol F, Durak C, et al. Evaluation of Continuous Renal Replacement Therapy and Therapeutic Plasma Exchange, in Severe Sepsis or Septic Shock in Critically Ill Children. *Medicina (Kaunas).* 2019;55(7): 350.
- [13] Sık G, Demirbuga A, Annayev A, Akcay A, Çıtak A, Öztürk G. Therapeutic plasma exchange in pediatric intensive care: Indications, results and complications. *Ther Apher Dial.* 2020;24(2): 221-229.
- [14] Atay G, Yazar H, Erdoğan S, Tuğrul HC, Işcan H, Kutlubay B. Therapeutic Plasma Exchange for Treating Pediatric Neurological Diseases. *Trends in Pediatrics.* 2022;3(2): 47-50.
- [15] Taylan C, Schaaf A, Dorn C, et al. Safety of Therapeutic Apheresis in Children and Adolescents. *Front Pediatr.* 2022;10: 850819. Published 2022 Apr 12.
- [16] Savransky A, Rubstein A, Rios MH, et al. Prognostic indicators of improvement with therapeutic plasma exchange in pediatric demyelination. *Neurology.* 2019;93(22): e2065-e2073.
- [17] Mörtzell Henriksson M, Newman E, Witt V, et al. Adverse events in apheresis: An update of the WAA registry data. *Transfus Apher Sci.* 2016;54(1): 2-15.




- [18] Gala-Błądzińska A, Mazur K, Dębiec A, Gargas K, Bartosik-Psujek H. Safety and tolerability of therapeutic plasma exchange in autoimmune neurological diseases - a retrospective single-centre analysis. *Neurol Neurochir Pol.* 2020;54(4): 344-349.
- [19] Duyu M, Turkozkan C. Therapeutic plasma exchange in the pediatric intensive care unit: A single-center 5-Year experience. *Transfus Apher Sci.* 2020;59(5): 102959.
- [20] Cortina G, Ojinaga V, Giner T, et al. Therapeutic plasma exchange in children: One center's experience. *J Clin Apher.* 2017;32(6): 494-500.
- [21] Lemaire A, Parquet N, Galicier L, et al. Plasma exchange in the intensive care unit: Technical aspects and complications. *J Clin Apher.* 2017;32(6):405-412.
- [22] Kalenderoğlu MD, Çomak E, Aksoy GK, et al. Therapeutic apheresis: is it safe in children with kidney disease? *Pediatr Nephrol.* 2024;39(8): 2451-2457.
- [23] Reddy SK, Jahan A, Chaturvedi S, Agarwal I. Plasma exchange for paediatric kidney disease- indications and outcomes: a single-centre experience. *Clin Kidney J.* 2015;8(6): 702-707.
- [24] Mokrzycki MH, Balogun RA. Therapeutic apheresis: a review of complications and recommendations for prevention and management. *J Clin Apher.* 2011;26(5): 243-248.
- [25] Gafoor VA, Jose J, Saifudheen K, Musthafa M. Plasmapheresis in neurological disorders: Experience from a tertiary care hospital in South India. *Ann Indian Acad Neurol.* 2015;18(1): 15-19.
- [26] Eyre M, Hachon Y, Barton C, Hemingway C, Lim M. Therapeutic plasma exchange in paediatric neurology: a critical review and proposed treatment algorithm. *Dev Med Child Neurol.* 2018;60(8): 765-779.
- [27] Michon B, Moghrabi A, Winikoff R, et al. Complications of apheresis in children. *Transfusion.* 2007;47(10): 1837-1842.
- [28] Runowski D, Prokurat S, Rubik J, Grenda R. Therapeutic Plasma Exchange in Pediatric Renal Transplantation Experience of One Decade and 389 Sessions. *Transplant Proc.* 2018;50(10): 3483-3486.

- 395 [29] Bobati SS, Naik KR. Therapeutic Plasma Exchange - An Emerging Treatment Modality in Patients  
396 with Neurologic and Non-Neurologic Diseases. *J Clin Diagn Res.* 2017;11(8): EC35-EC37.
- 397 [30] Webb TN, Bell J, Griffin R, Dill L, Gurosky C, Askenazi D. Retrospective analysis comparing  
398 complication rates of centrifuge vs membrane-based therapeutic plasma exchange in the pediatric  
399 population. *J Clin Apher.* 2022;37(3):263-272. doi:10.1002/jca.21969.
- 400 [31] Kielstein JT, Hafer C, Zimbudzi E, Hawes S. A Change for Better Exchange – From Membrane  
401 Therapeutic Plasma Exchange to Centrifugal Therapeutic Plasma Exchange. *European Medical*  
402 *Journal.* 2020 Mar;8(Suppl 1):2-10.
- 403 [32] Meyer EK, Wong EC. Pediatric Therapeutic Apheresis: A Critical Appraisal of Evidence.  
404 *Transfusion Medicine Reviews.* 2016 Oct;30(4): 217-222.

# Therapeutic Plasma Exchange in Pediatrics



**178 patients**  
740 procedures

 Predominantly femoral access

**118.5 hours**  
Median catheter duration

## Complication rates



**20.8%**  
Technical complications



**4.2%**  
Clinical complications

Technical complications were **5x** more frequent than clinical

## Risk factors



Each day of catheter use  
+ **3 %** risk



Lower age  
+ **8.9%** risk of clinical complications per year



+ **21.6%** clinical complication risk

## Tables

Table 1. Epidemiological data of the studied population, doses of supplements used, and technical data of TPE procedures: NE subgroup (81 patients; 107 sessions; 360 TPEs); non-NE subgroup (97 patients, 107 sessions, 380 TPEs); entire studied population (178 patients, 214 sessions; 740 TPEs). Data are presented as medians with interquartile ranges (IQR); p-value – Wilcoxon test. \*Data pertain to the first TPE in a given session \*\*surrogate for catheter usage time; for single TPE sessions, the time corresponds to the duration of the specific TPE. ACT – activated clotting time; QB – blood flow velocity; HES - Hydroxyethyl Starch; FFP – Fresh Frozen Plasma.

Characteristics of studied population and TPE procedures	NE subgroup	Non-NE subgroup	p - value	All studied population
Age [months]	166.0 (104.0)	120.00 (104.38)	NS	138.00 (105.00)
Body mass [kg]	44.5 (30.0)	32.00 (32.40)	NS	38.00 (32.00)
Body mass [percentiles]	29.0 (58.0)	50.00 (58.00)	NS	39.00 (60.00)
Height [cm]	160.00 (43.00)	138.25 (45.00)	0.043	146.00 (47.50)
Height [percentiles]	45.00 (42.00)	40.00 (60.75)	NS	40.50 (52.00)
<b>Dosages of supplements</b>				
FFP [ml/kg]	33.6 (17.7)	30.00 (24.88)	0.008	32.00 (20.20)
5% Albumin [ml/kg]	38.5 (28.2)	33.33 (23.52)	0.012	36.23 (26.72)
6% HES [ml/kg]	25.6 (15.6)	25.42 (13.34)	NS	25.53 (15.67)
Crystalloid fluids/ Ringer [ml/kg]	12.5 (8.3)	10.42 (7.55)	0.009	11.63 (8.99)
Total exchanged plasma volume [ml/kg]	73.1 (23.1)	70.16 (28.31)	0.025	72.12 (25.81)
FFP / total exchanged plasma volume	0.5 (0.1)	0.44 (0.22)	NS	0.46 (0.17)
FFP /5% Albumin	1.1 (0.8)	1.00 (0.85)	0.002	1.00 (0.86)
FFP/ 6% HES	1.1 (0.7)	1.08 (0.80)	NS	1.10 (0.70)
20% Albumin [ml/kg]	2.1 (0.9)	1.75 (1.35)	0.056	2.13 (1.23)
<b>Technical aspects of procedures</b>				
QB [ml/kg/min]	2.00 (1.3)	2.17 (1.52)	0.022	2.12 (1.48)
Duration of TPE procedure [min]	155.00 (90.00)	150.00 (80.00)	NS	155.00 (90.00)
Duration of TPE session [hours]** (all catheters)	145.00 (184.50)	49.00 (224.44)	0.027	118.50 (215.00)
Duration of TPE session [hours]** (excluding permanent catheters)	143 [168.8]	46.5 [188.7]	0.003	99 [192.9]
Supplement exchange flow rate [ml/kg/h]	28.3 (21.6)	28.49 (20.68)	NS	28.37 (21.11)
Heparin initial dose [mg/kg]	0.30 (0.29)	0.27 (0.25)	<0.001	0.29 (0.29)
Dosage of calcium supplementation [ml/kg] (only for TPE with FFP)	1.64 (0.81)	1.45 (1.04)	0.003	1.54 (0.92)

Table 2. Catheter-related technical and clinical complications in NE and Non-NE populations. NE subgroup (81 patients; 107 sessions; 360 TPEs); non-NE subgroup (97 patients, 107 sessions, 380 TPEs); entire studied population (178 patients, 214 sessions; 740 TPEs). Data are presented as the number of events and percentages; p-value – one-sided Fisher's exact test; AEs – adverse events.

Catheter-related technical complications	NE subgroup	Non-NE subgroup	p - value	All studied population
<b>No. of TPE with technical complication (% vs No. of TPE)</b>	106 (29.4%)	48 (12.6%)	<0.001	154 (20.8%)
<b>No. of particular technical AE (% vs No. of TPE)</b>				
Insufficient blood intake – malfunction in the arterial part of the catheter	89 (24.7%)	28 (7.3%)	<0.001	117 (15.8%)
Reversed lines	68 (18.9%)	35 (9.2%)	<0.001	103 (13.9%)
Insufficient blood return – malfunction in the venous part of the catheter	5 (1.4%)	5 (1.3%)	NS	10 (1.4%)
Thrombosis within the catheter	1 (0.3%)	1 (0.3%)	NS	2 (0.3%)
Spontaneous catheter repositioning	0	1 (0.3%)	NS	1 (0.1%)
Catheter leakage	0	1 (0.3%)	NS	1 (0.1%)

<b>No. of all technical AEs (% vs No. of TPE)</b>	163 (45.3%)	71 (18.7%)	<0.001	234 (31.6%)
<b>Catheter-related clinical complications</b>				
<b>No. of TPE with clinical complication (% vs No. of TPE)</b>	14 (3.9%)	17 (4.5%)	NS	31 (4.2%)
<b>No. of particular clinical AE (% vs No. of TPE)</b>				
Thrombosis in the vessel with the catheter in place	6 (1.7%)	5 (1.3%)	NS	11 (1.5%)
Bleeding at the implantation site	2 (0.6%)	8 (2.1%)	NS	10 (1.4%)
Catheter's exit site infection	3 (0.8%)	3 (0.8%)	NS	6 (0.8%)
Pain at the catheter implantation site	4 (1.1%)	2 (0.5%)	NS	6 (0.8%)
Edema of the extremity with the catheter	1 (0.3%)	1 (0.3%)	NS	2 (0.3%)
<b>No. of all catheter-related clinical AEs (% vs No. of TPE)</b>	16 (4.4%)	19 (5%)	NS	35 (4.7%)

Table 3. Medical interventions required to catheter-related technical and clinical complications in NE and Non-NE Populations. NE subgroup (81 patients; 107 sessions; 360 TPEs); non-NE subgroup (97 patients, 107 sessions, 380 TPEs); entire studied population (178 patients, 214 sessions; 740 TPEs). Data are presented as the number of events and percentages; p-value – one-sided Fisher's exact test; AEs – adverse events; rTPA – recombinant tissue-type plasminogen activator.

<b>Type of intervention (% vs No. of TPE)</b>	<b>NE subgroup</b>	<b>Non-NE subgroup</b>	<b>p - value</b>	<b>All studied population</b>
Low Molecular Weight Heparin	8 (2.2%)	5 (1.3%)	NS	13 (1.8%)
Surgical pressure dressing	1 (0.3%)	8 (2.1%)	0.023	9 (1.2%)
Topical antibiotic	3 (0.8%)	3 (0.8%)	NS	6 (0.8%)
Additional peripheral vascular access	0	5 (1.3%)	0.035	5 (0.7%)
Urokinase / Turolock (administered via catheter)	3 (0.8%)	1 (0.3%)	NS	4 (0.5%)
rTPA / Actylise (intravenous)	2 (0.6%)	1 (0.3%)	NS	4 (0.5%)
Flushing the catheter with saline	3 (0.8%)	1 (0.3%)	NS	4 (0.5%)
Administration of analgesic medication	2 (0.6%)	2 (0.5%)	NS	4 (0.5%)
Changing the position of the extremity	2 (0.6%)	1 (0.3%)	NS	3 (0.4%)
Removal of a single suture and adjustment of the catheter	2 (0.6%)	0	NS	2 (0.3%)
Cooling dressing	2 (0.6%)	0	NS	2 (0.3%)
Transfer of the patient to the ICU	1 (0.3%)	0	NS	1 (0.1%)
Catheter replacement	1 (0.3%)	0	NS	1 (0.1%)
Surgical vessel suturing under general anesthesia	1 (0.3%)	0	NS	1 (0.1%)
<b>Total No. of interventions (% vs. No. of TPEs)</b>	31 (8.6%)	27 (7.1%)	NS	58 (7.8%)
<b>Total No. of patients with catheter-related interventions (% vs. No. of patients)</b>	24 (29.6%)	17 (17.5%)	0.019	41 (23%)

Table 4. Technical difficulties related to the filter, machine, and causes of premature termination of the TPE in NE and non-NE populations. NE subgroup (81 patients; 107 sessions; 360 TPEs); non-NE subgroup (97 patients, 107 sessions, 380 TPEs); entire studied population (178 patients, 214 sessions; 740 TPEs). Data are presented as the number of events and percentages; p-value – one-sided Fisher's exact test; AEs – adverse events.

<b>Technical AEs related to filters and machine</b>	<b>NE subgroup</b>	<b>Non-NE subgroup</b>	<b>p - value</b>	<b>All studied population</b>
<b>No. of TPE with technical AEs related to filters (% vs No. of TPE)</b>	24 (6.7%)	14 (3.7%)	0.047	38 (5.1%)
<b>No. of particular AE related to filters (% vs No. of TPE)</b>				
Filter clotting	16 (4.4%)	13 (3.4%)	NS	29 (3.9%)
Filter replacement	16 (4.4%)	8 (2.1%)	0.056	24 (3.2%)
Increased pressure on the filter without clotting	6 (1.6%)	4 (1.1%)	NS	10 (1.3%)

Capillary rupture in the filter	4 (1.1%)	0	0.056	4 (0.5%)
<b>Total No. of filter-related AEs (% vs No. of TPE)</b>	<b>42 (11.7%)</b>	<b>25 (6.6%)</b>	<b>0.011</b>	<b>67 (9.1%)</b>
<b>No. of TPE with technical AEs related to machines (% vs No. of TPE)</b>	<b>3 (0.8%)</b>	<b>0</b>	<b>NS</b>	<b>3 (0.3%)</b>
<b>No. of TPE prematurely terminated (% vs No. of TPE)</b>	<b>16 (4.4%)</b>	<b>6 (1.6%)</b>	<b>0.018</b>	<b>22 (3%)</b>
<b>No. of particular AEs related to prematurely terminated TPE (% vs No. of TPE)</b>				
Filter malfunction	12 (3.3%)	6 (1.6%)	NS	18 (2.4%)
CVC malfunction	7 (1.9%)	0	0.006	7 (1%)
CVC-related clinical complications	1 (0.3%)	0	NS	1 (0.1%)
Machine failure	1 (0.3%)	0	NS	1 (0.1%)
Damaged FFP bag	1 (0.3%)	0	NS	1 (0.1%)
<b>Total No. of AEs related to prematurely terminated TPE (% vs No. of TPE)</b>	<b>22 (6.1%)</b>	<b>6 (1.6%)</b>	<b>0.001</b>	<b>28 (3.8%)</b>

Table 5. Catheter-related technical and clinical complications in FFP1 and FFP0 populations. FFP1 population: 609 TPEs; FFP0 population: 131 TPEs. Data are presented as the number of events and percentages; p-value – one-sided Fisher’s exact test; AEs – adverse events.

<b>Catheter-related technical complications</b>	<b>FFP1 subgroup</b>	<b>FFP0 subgroup</b>	<b>p value</b>
<b>No. of TPE with technical complication (% vs No. of TPE)</b>	<b>140 (23%)</b>	<b>14 (10.7%)</b>	<b>0.001</b>
<b>No. of particular technical AE (% vs No. of TPE)</b>			
Insufficient blood intake – malfunction in the arterial part of the catheter	89 (14.6%)	28 (21.4%)	0.040
Reversed lines	81 (13.3%)	22 (16.8%)	NS
Insufficient blood return – malfunction in the venous part of the catheter	8 (1.3%)	2 (1.5%)	NS
Thrombosis within the catheter	2 (0.3%)	0	NS
Spontaneous catheter repositioning	1 (0.2%)	0	NS
Catheter leakage	1 (0.2%)	0	NS
<b>No. of all technical AEs (% vs No. of TPE)</b>	<b>182 (29.9%)</b>	<b>52 (39.7%)</b>	<b>0.020</b>
<b>Catheter-related clinical complications</b>			
<b>No. of TPE with clinical complication (% vs No. of TPE)</b>	<b>30 (4.9%)</b>	<b>1 (0.8%)</b>	<b>0.017</b>
<b>No. of particular clinical AE (% vs No. of TPE)</b>			
Thrombosis in the vessel with the catheter in place	10 (1.6%)	1 (0.8%)	NS
Bleeding at the CVC implantation site	10 (1.6%)	0	NS
Catheter’s exit site infection	6 (1%)	0	NS
Pain at the catheter implantation site	6 (1%)	0	NS
Edema of the extremity with the catheter	2 (0.3%)	0	NS
<b>No. of all catheter-related clinical AEs (% vs No. of TPE)</b>	<b>34 (5.6%)</b>	<b>1 (0.8%)</b>	<b>0.008</b>

Table 6. Technical difficulties related to the filter, machine, and causes of premature termination of the TPE in FFP1 and FFP0 populations. FFP1 population: 609 TPEs; FFP0 population: 131 TPEs. Data are presented as the number of events and percentages; p-value – one-sided Fisher’s exact test; AEs – adverse events.

<b>Technical AEs related to filters and machine</b>	<b>FFP1 subgroup</b>	<b>FFP0 subgroup</b>	<b>p value</b>
<b>No. of TPE with technical AEs related to filters (% vs No. of TPE)</b>	<b>20 (3.3%)</b>	<b>18 (13.7%)</b>	<b>&lt;0.001</b>
<b>No. of particular AE related to filters (% vs No. of TPE)</b>			
Filter clotting	14 (2.3%)	15 (11.4%)	<0.001

Filter replacement	12 (2%)	12 (9.2%)	<0.001
Increased pressure on the filter without clotting	7 (1.1%)	3 (2.3%)	NS
Capillary rupture in the filter	3 (0.5%)	1 (0.8%)	NS
<b>Total No. of filter-related AEs (% vs No. of TPE)</b>	<b>36 (5.9%)</b>	<b>31 (23.7%)</b>	<b>&lt;0.001</b>
<b>No. of TPE with technical AEs related to machines (% vs No. of TPE)</b>	<b>2 (0.3%)</b>	<b>1 (0.8%)</b>	<b>NS</b>
<b>No. of TPE prematurely terminated (% vs No. of TPE)</b>	<b>12 (2%)</b>	<b>10 (7.6%)</b>	<b>0.002</b>
<b>No. of particular AEs related to prematurely terminated TPE (% vs No. of TPE)</b>			
Filter malfunction	11 (1.7%)	7 (5.3%)	0.026
CVC malfunction	3 (0.5%)	4 (3.1%)	0.021
CVC-related clinical complications of TPE	1 (0.2%)	0	NS
Machine failure	0	1 (0.8%)	NS
Damaged FFP bag	1 (0.2%)	0	NS
<b>Total No. of AEs related to prematurely terminated TPE (% vs No. of TPE)</b>	<b>16 (2.6%)</b>	<b>12 (9.2%)</b>	<b>0.001</b>

Table 7. Multivariable logistic regression evaluating the influence of selected variables on occurrence of a technical and clinical complication during a given TPE divided into all and only acute catheters; OR values for selected factors; CI95 intervals for OR values; p-value. OR - odds ratio; CI – confidence interval; independent variables: \*presented as the number of FFP1 TPEs within a given session;

	Age [years]		Use of FFP (1/0) *		Affiliation with the NE subgroup (1/0)		Days since catheter placement	
	OR (CI95)	p-value	OR (CI95)	p-value	OR (CI95)	p-value	OR (CI95)	p-value
All catheters	---			---	2.396 (1.347-4.263)	0.003	1.030 (1.003-1.057)	0.029
Technical Complication								
Clinical Complication	0.911 (0.840-0.988)	0.024	1.216 (1.016-1.456)	0.033	---		---	
Acute catheters only	OR (CI95)	p-value	OR (CI95)	p-value	OR (CI95)	p-value	OR (CI95)	p-value
Technical Complication	---			---	2.090 (1.175-3.718)	0.012	1.060 (1.015-1.108)	0.008
Clinical Complication	0.909 (0.839-0.986)	0.021	1.208 (1.012-1.442)	0.037	---		---	