

Continuous glucose monitoring (CGM) systems in Poland, Europe, and beyond in 2025: principles, available technologies, and future perspectives

Oskar Kublin^{1*}, Dorota Filipowicz², Mariusz Stępień¹

¹Department of Family Medicine, Internal Diseases and Social Pharmacology, Medical University of Lodz, Poland

²Department of Endocrinology, Metabolism and Internal Medicine, Poznan University of Medical Sciences, Poznan, Poland

Submitted: 13 August 2025; **Accepted:** 17 February 2026

Online publication: 18 August 2026

Arch Med Sci 2026; 22 (4): 2250–2265

DOI: <https://doi.org/10.5114/aoms/218227>

Copyright © 2026 Termedia & Banach

***Corresponding author:**

Oskar Kublin, MD, PhD
Department of Family Medicine, Internal Diseases and Social Pharmacology
Medical University of Lodz
Żeromskiego 90-549
Lodz, Polska
Phone: +48 42 639 37 50
E-mail: oskar.kublin@umed.lodz.pl

Abstract

In recent years, continuous glucose monitoring (CGM) systems have become increasingly common and more widely available for people with diabetes. A variety of devices appearing on the market, together with their increasing integration with other tools (insulin pumps, watches, and smartphone applications), require medical personnel to have a solid basic knowledge of CGMs, continuous follow-up, updated expertise extending beyond the local market, and a future-oriented perspective. Additionally, low-quality products without proper evaluation, which do not meet the standards established by diabetes societies, become available to patients. To provide an extended background on CGM technology principles, basic knowledge of results presentation, and interpretation according to recommendations, this review presents the most common devices, along with a cost-effectiveness analysis and a critical assessment. Furthermore, future perspectives are discussed, including AI technologies and modern solutions to enhance user comfort.

Key words: continuous glucose monitoring, diabetes type 1, advanced hybrid closed-loop, diabetes technology, insulin.

Introduction

The world's first commercial three-day continuous glucose monitoring (CGM) system was approved by the U.S. Food and Drug Administration (FDA) and obtained Conformité Européenne (CE) marking around 2000 [1]. Initially, unlike other available methods for measuring glucose in interstitial fluid, systems used flash glucose monitoring (FGM) technology. It had no alarms and required the reader to approach the sensor closely to obtain a result. Today, all available systems operate on the principle of real-time continuous glucose monitoring (rtCGM). According to the manufacturer of the most common device, FreeStyle Libre is used worldwide by more than 5 million people in 60 countries [2]. Moreover, the recently released products are intended not only for insulin-dependent people with diabetes but also for those on oral antidiabetic medications and those seeking lifestyle interventions and a better understanding of their metabolic responses [1]. Less than a decade ago, the first device appeared on the Polish market, and since then, CGMs have become widespread, especially after the introduction of partial re-

imbursement in 2018. Over time, different groups of patients became eligible for reimbursement, for which various physicians (i.e., general practitioners, gynecologists, and ophthalmologists) and nurses are authorized to prescribe CGMs [3]. Experts in diabetic care emphasize the need for closer scrutiny of CGM quality [4, 5]. To systematize knowledge, improve understanding of reliable devices, and enhance decision-making for health providers and people with diabetes, this review article describes the basic principles of CGM systems and compares currently available CGM devices in Europe, spotlighting the latest technologies and future perspectives in glucose management.

Glucose monitoring

Glucometers

Although a wide range of CGM systems is available on the market, glucometers remain the most commonly used devices for monitoring blood glucose. Patients can select the most suitable model from among various devices. Glucometers differ slightly, for example, in size, display quality, or additional features [6]. Some systems can also send results to a pump or a dedicated smartphone application. All currently available glucometers should meet ISO EN 15197:2015 (International Organization for Standardization European Norm) standards, allowing for a 15% measurement error [7]. Regardless of the model, the patient has to make a puncture to obtain a blood sample.

CGM systems

Unlike glucometers, CGM systems measure glucose painlessly in the interstitial fluid. The patient receives the current glucose result, a predicted glucose level change, and has access to previous measurements (past, present, and future) (Figure 1). The location of the measurement is crucial for the accurate interpretation of the results. The change in glucose concentration in the interstitial fluid is delayed compared to the concentration in the blood [8]. This delay is called “lag time”, and current CGM systems use predictive algorithms to reduce the difference [9–11]. The user should consider this information when making informed decisions about their therapy. Depending on the CGM system, the patient may use a variety of alarms. The basic ones are glycaemic range alarms, which notify them when hypoglycemia or hyperglycemia occurs. More sophisticated systems display a rapid increase or decrease in glucose level, and are equipped with predictive algorithms for hypoglycemia and hyperglycemia. Depending on the device, alarms can be personalized to be either audible or vibrating.

Measurement accuracy in CGM systems is assessed by the mean absolute relative difference (MARD) parameter. A lower MARD value indicates higher accuracy of the measurement. This value should be below 10 to dose insulin based on measurements from a CGM sensor [12]. Currently, all systems available on the Polish market meet this requirement. Table I shows differences between measurements made with a glucometer and a CGM system.

Result interpretation

Compared to self-monitoring of blood glucose (SMBG), rtCGM systems measure glucose more frequently, typically every 1–5 min, which means the systems generate between 288 and 1440 measurements per day. Results obtained from the sensor are automatically transmitted via Bluetooth to a receiving device; these days, it is usually an application on a smartphone or a smartwatch, or a dedicated receiver. Then, upon the user's consent, they can be transmitted to other people (caregivers or medical personnel) via the internet. Reports are generated from the obtained results. The current standard is the ambulatory glucose profile (AGP) report, established by an expert consensus in 2019 [13–16].

Ten parameters should be considered when evaluating CGM results: 1) duration of use of the device – 14 days is recommended; 2) duration of




Figure 1. Result from CGM. Past, present, future

Table I. Comparison of SMBG and CGM measurements

Parameter	SMBG	CGM
Measurement	Spot, painful	Continuous, painless
Obtained result	Current blood glucose	Current blood glucose, change trend, last period results
Alarm	Unavailable	Available, customizable
Complications	Fingertip trauma	Skin lesions around the sensor
Cost	Lower	Higher
Availability	Higher	Lower
Number of available devices	Higher	Lower
Measurement error	Up to 15% of ISO standard	Up to 10% (MARD of currently available devices)
Cost reimbursement	A larger group of patients – all people with diabetes	A smaller group of patients – some people with diabetes

CGM – continuous glucose monitoring, ISO – International Organization for Standardization, MARD – mean absolute relative difference, SMBG – self-measurements of blood glucose.

rtCGM sensor activity – this should be a minimum of 70% of 14 days; 3) average glycemia; 4) glucose management indicator (GMI) – recommended values as for glycated hemoglobin (HbA1c); 5) glycemic variability (%CV) – this should be $\leq 36\%$.

The other parameters demonstrate the percentage of time within specific glycemic ranges. They can be interpreted as follows: (6) time above range (TAR): % of readings and time > 250 mg/dl (> 13.9 mmol/l) – level 2 – very high; 7) time above range (TAR): % of readings and time 181–250 mg/dl (10.1–13.9 mmol/l) – level 1 – high; 8) time in range (TIR): % of readings and time 70–180 mg/dl (3.9–10.0 mmol/l); 9) time below range (TBR): % of readings and time 54–69 mg/dl (3.0–3.8 mmol/l) – level 1 – low; 10) time below range (TBR): % of readings and time < 54 mg/dl (< 3.0 mmol/l) – level 2 – very low.

Due to the measurement of glycemia every few minutes, glycemic variability is now accurately calculated. Importantly, %CV is a risk factor for complications, and when higher than 36%, increases the likelihood of hypoglycemia [17, 18]. GMI is another new parameter [19]. Unlike HbA1c, the value of the GMI parameter is calculated based on a shorter period of time, and the duration of sensor activity is crucial for obtaining reliable results. Opposite to HbA1c, GMI is not dependent on factors such as anemia, administered medication, or erythrocyte lifespan [20]. The TIR value established in the 2019 consensus refers to a glucose range of 70–180 mg/dl (3.9–10.0 mmol/l) for the general population and 63–140 mg/dl (3.5–7.8 mmol/l) for pregnant women with diabetes. Time in tight range (TIR), i.e., 70–140 mg/dl (3.9–7.8 mmol/l), is a new parameter that is increasingly appearing in publications [21, 22]. Time in the tight pregnancy range, with a value of 63–120 mg/dl (3.5–6.7 mmol/l), has also been proposed

[23]. Therapeutic targets for different glycemic ranges are discussed in the following paragraph.

AGP reports generated by software of individual CGM systems should meet the current standard and include all parameters as presented in detail in the 2019 consensus [24]. The graphical representation of results varies slightly between devices made by different manufacturers. In addition, reports from rtCGM systems connected to insulin pumps may include data on insulin pump therapy.

Polish Diabetes Association (PTD) recommendations

Therapeutic targets recommended by the PTD in 2025 are consistent with those proposed in the previous report [15, 25]. According to the PTD, CGM systems are the preferred method of monitoring glucose in patients with type 1 diabetes [26]. Patients using CGM systems in combination with advanced hybrid closed-loop (AHCL) systems will benefit most. To increase the safety of therapy, improve patient comfort, and enhance the quality of care, the PTD recommends the use of CGM systems for all patients with diabetes treated with intensive insulin therapy or continuous subcutaneous insulin infusion (CSII) [27].

TIR values for patients with type 1 and type 2 diabetes, as well as older adults, should be between 70 and 180 mg/dl (3.9–10.0 mmol/l). Older adults and patients at high risk of hypoglycemia should demonstrate the above values for at least 50% of the day. Other patients are expected to have these values for at least 70% of the day (16 h and 48 min). For pregnant women with diabetes, the target range for TIR should be 63–140 mg/dl (3.5–7.8 mmol/l) for at least 70% of the day. However, the group of children and adolescents should exhibit the following values: TIR $\geq 80\%$, CV $< 36\%$, and HbA1c $\leq 6.5\%$ (47.5 mmol/mol) [28]. During

remission and during insulin pump treatment with AHCL, TITR and a reduced CV < 33% may be indicated.

TBR < 54 mg/dl (< 3.0 mmol/l) in the group of patients with type 1 and type 2 diabetes and pregnant women with diabetes should be < 1% (15 min). The group of pregnant women should demonstrate TBR – level 1 < 63 mg/dl (< 3.5 mmol/l), and a value of < 4% (1 h) is recommended. In people with type 1 and type 2 diabetes, the TBR level 1 with a value < 70 mg/dl (< 3.9 mmol/l) should not exceed 4% of the day. Additionally, TBR in older patients should not exceed 1% daily at a value < 70 mg/dl (< 3.9 mmol/l).

The TAR value for pregnant women with diabetes is set at 140 mg/dl (7.8 mmol/l), and it should stay within this range for less than 25% of the day (6 h). For other patients, two threshold values for TAR have been established. Level 1, defined as 181–250 mg/dl (10.1–13.9 mmol/l), is allowed for up to 50% of the day (12 h) in older patients, whereas in younger patients, the level is recommended for up to 25% of the day (6 h). Level 2, i.e., > 250 mg/dl (> 13.9 mmol/l), is allowed for up to 10% of the day (2 h 24 min) in older adults, whereas in other patients, it should be limited to 5% of the day (1 h 12 min).

CGM systems

Basic available CGM systems can be categorized by: a) duration of use into short-term and long-term, b) type of implantation into transcutaneous and subcutaneous, c) measurement method into enzymatic and fluorescence.

Additionally, the various devices differ in whether they require the sensor to be connected to a transmitter, which then transmits the results to the receiving device. The sensor can also be integrated into the transmitter. Transmitters can have a specific operating time and may require periodic recharging. Most sensors are sold with a single-useserter. Some other sensors require the purchase of a dedicatedserter for installation. The location for sensor placement, the minimum age limit for using such a device, and recommendations regarding its use by pregnant women with diabetes can vary among different manufacturers. Nowadays, almost all sensors are factory-calibrated; some additionally can be calibrated by the user if the manufacturer does not prohibit calibration.

After insertion, the sensor requires a warm-up period of 30–120 min. A subcutaneously implanted sensor is unusual because it becomes effective after 24 min. Transdermal sensors leave a thin electrode in the interstitial fluid during implantation. The electrode is coated with the enzyme glucose oxidase. This enzyme, when combined with glucose, is oxidized. The release of free electrons

accompanies this process. The sensor measures the current voltage, which proportionally corresponds to the concentration of glucose in the interstitial fluid. Algorithms then process the obtained data, and the user is given a glucose concentration result every 1–5 min.

CGM systems available on the market are described below. They are listed in order from the shortest to the longest running time. Comparative data (key features, pump integrability, clinical parameters) of various CGMs can be found in Table II. The CGMs are presented with their absolute MARD values and pump integration possibilities (Figure 2) and application method (Figure 3). A cost-effectiveness analysis with a detailed description of reimbursement in Poland is presented in Supplementary Material I. The reimbursement of the FreeStyle Libre system in selected European countries is additionally described. Applications required for CGM systems reading and sharing are presented in Table III.

Medtronic

Medtronic offers several types of rtCGM systems. Currently produced sensors are either a continuation of an earlier product, Guardian 4, or a new product, Simplerla [29, 30]. The Guardian 4 sensor is implanted using a dedicated reusableserter. The patient must also purchase a transmitter and charger [31]. The transmitter must be charged each time the sensor is replaced. Sensor running time is 7 days. The manufacturer gives a 12-month warranty on the transmitter. The Guardian 4 sensor can be used in combination with a MiniMed 780G insulin pump or in the SMART CGM system [29, 32]. The MiniMed Mobile application is used to operate both the solutions [33, 34]. However, when the sensor is connected to the pump, results are transmitted to the application via the pump, so it is not possible to use the rtCGM system without the pump. Results can also be transmitted to the user's carers [35, 36]. Additionally, measurements can be shared with the user's CareLink Personal account, where reports can be generated [37]. Medical staff can, upon the patient's consent, receive their results [38].

The Simplerla sensor does not require a transmitter, as it is already integrated [30]. The sensor is implanted with a single-useserter. This device also works for 7 days. Results are sent to the pump or application [39, 40]. The system can also work with an InPen smart injector (which the patient must purchase) and the InPen application [41–43]. A bolus calculator is then available in the application. The patient can control insulin injections and will receive predictive alerts. Both Simplerla and Guardian 4 can be integrated with the 780G pump, forming a closed loop [32]. In this device,

Table II. CGM systems available in 2025 – expanded to include pump integrations (AHCL readiness), key device features, and clinical parameters (age indications, pregnancy compatibility, wear duration, warm-up period, calibration requirements, and insertion sites)

Manufacturer/ Distribution	Sensor/Transmitter required	Running time [days]	Registration from [year of age]	Possibility of calibra- tion	Warm- up time [h]	Registration pregnant	Implantation site	Integration with pump/pen/other
Medtronic	Guardian 4/transmitter 12-month	7	7	Optional	2	No	> 18 years – arm, abdomen 7–17 – additionally buttock	MiniMed 780G/InPen/Apple Watch
	Simplera	7	2	Optional	2	No	> 18 years – arm 7–17 – additionally buttock	MiniMed 780G/InPen/Apple Watch
Dexcom/ Proglukemia	G6/transmitter 3-month	10	10	Optional	2	Yes	> 2 years abdomen; children 2–17 – additionally buttock	YpsoPump/Garmin devices
	One+	10 (+0.5)	2	Optional	0.5	Yes	> 2 years abdomen and arm; children 2–6 – additionally above the buttock	No
	G7	10 (+.5)	2	Optional	0.5	Yes	> 2 years abdomen and arm; children 2–6 – additionally above the buttock	YpsoPump (no in PL)/Garmin devices/Apple Watch
Abbott	Libre 2	14	4	No	1	Yes	Arm	No
	Libre 3	14	4	No	1	Yes	Arm	YpsoPump (no in PL)
	Libre 2 Plus	15	2	No	1	Yes	Arm	Omnipod 5
	Libre 3 Plus	15	2	No	1	Yes	Arm	Ypso Pump/Beta Bionics iLet Bionics Pancreas/(AID) System Powered by Tidepool
	Instinct	15	7	No	1	No	Arm	MiniMed 780G
Roche	Accu-Chek SmartGuide	14	18	Required on day 1	1	Yes	Arm	Apple Watch
Medtrum/ Diagnosis	S9 CGM Sensor TouchCare/transmitter 12-month	14	2	Optional	0.5	Yes	Adults – abdomen and arm; children – abdomen and buttock	No
SiSensing / Diagnosis	Sibionics GS1	14	3	No	2	No	Arm	No
Sinocare	iCan I3	15	18	No	2	No	Abdomen	No
i-SENS /Bioton	CareSens Air	15	18	1× daily	0.5	No	Arm	No
Synoptis	LynX	15	18	Optional	1	No	Arm	Equil pump (no in PL)
Ascensia	Eversense E3/ transmitter 12-month	180	18	2× daily on days 1–21; 1× daily from day 21	24	No	Arm – subcutaneously	Apple Watch




Figure 2. Timeline of CE-marked continuous glucose monitoring (CGM) systems introduced in Europe between 2018 and September 2025. The timeline indicates the year of CE mark approval, with the month (Roman numerals) displayed before each brand name. An insulin pump icon adjacent to the CGM label denotes the possibility of integration into automated closed-loop systems. The yellow circles indicate the mean absolute relative difference (MARD) values provided by the manufacturers. Ascensia Eversense E3 is the only implantable, clinician-inserted subcutaneous sensor with up to 6 months of wear, whereas all other CGM devices are self-inserted transcutaneous sensors with wear durations ranging from 7 to 15 days. The number of available devices is increasing




Figure 3. Classification of commonly used CGM systems by application method: (bottom) surgically implantable, long-term subcutaneous sensors and (top) user-applied, transcutaneous sensors inserted with a dedicated inserter

insulin is automatically dosed based on the current glucose level, and the patient only needs to enter the amount of carbohydrates consumed before a meal. The system stops insulin supply when glucose concentration drops. If it rises, the device increases the base or makes autocorrections. The 780G pump is one of three available AHCLs on the Polish market. It has been confirmed that this device improves metabolic control of diabetes [44].

Dexcom

This is another company whose products have been sold for many years. Currently, G6 is still sold, and more recent versions of One+ and G7 are available. A 10-day G6 sensor connects to a transmitter (whose running time is 3 months). All Dexcom sensors come with a single-use applicator. The G6 sensor is compatible with an application or dedicated receiver [45]. The patient can share their results with carers [46]. The user can generate reports and share them with medical staff [47]. The Dexcom G6 system can be connected to the YpsoPump via the mylife CamAPS FX application, creating a closed loop [48, 49]. This is the second solution of that kind available on the Polish market. Its positive effect in the treatment of diabetes has also been confirmed [50].

Dexcom's more recent sensors, One+ and G7, look similar. They both have an integrated transmitter and can work for 10.5 days. They come with a single-use serter. Results are sent to a smartphone application or to a dedicated receiver [51–53]. There are plans to extend the G7 running time to 15 days [54]. Notably, Dexcom can be connected to Garmin devices. Additionally, the “direct to watch” function enables the use of the G7 sensor without a phone by connecting the sensor directly to an Apple Watch [55, 56]. The functions above can be handy for sportspeople.

Libre

The second generation of the FreeStyle Libre system has replaced the previous version introduced almost a decade ago. The Libre 2 version is already a complete rtCGM system; a 14-day sensor is supplied with a single-use serter. To activate the sensor, it must be scanned using a smartphone with near field communication (NFC). The user can receive results in an application or a dedicated reader [57, 58]. They are sent automatically via Bluetooth, but can also be obtained by scanning the sensor. This is also recommended in the event of a signal loss. We can complete the records obtained during the last 8 h. Phone appli-

Table III. CGM applications

Manufacturer/ Distribution	Sensor	User application	Follower application	Medical personnel application or personal account
Medtronic	Guardian 4	MiniMed Mobile	CareLink Connect	CareLink Professional, CareLink Personal
	Simplera	Medtronic Simplera	CareLink Connect	CareLink Professional, CareLink Personal
Dexcom/ Proglukemia	G6	Dexcom G6	Dexcom Follow	Dexcom Clarity
	One+	Dexcom ONE+	Dexcom Follow	Dexcom Clarity
	G7	Dexcom G7	Dexcom Follow	Dexcom Clarity
Abbott	Libre 2	LibreLink	LibreLinkUp	LibreView
	Libre 3	Libre 3	LibreLinkUp	LibreView
	Libre 2 Plus	LibreView	LibreLinkUp	LibreView
	Libre 3 Plus	Libre 3	LibreLinkUp	LibreView
Roche	Accu-Chek SmartGuide	SmartGuide – Accu-Chek CGM, Predict - ACCU-CHEK SmartGuide		Accu-Chek Care
Medtrum/ Diagnosis	S9 CGM Sensor TouchCare	EasySense	EasyFollow	EasyView PRO and EasyView Personal
SiSensing/ Diagnosis	Sibionics GS1	Sibionics		
Sinocare	iCan I3	iCan CGM	iCan Reach	iCan Review
i-SENS/Bioton	CareSens Air	CareSens Air	Sens365	Sens365 platform
Synoptis	LynX	LinX CGM		
Ascensia	Eversense E3	Eversense CGM	Eversense NOW	Eversense DMS

cation users can send their results to their therapy partners [59]. Patients can generate reports and communicate with medical staff through a dedicated application [60]. A more recent version of the FreeStyle Libre 3, currently unavailable in Poland, is slightly smaller and can be connected to the YpsoPump to form a closed loop [61, 62]. The Libre 3 sensor also works for 14 days and has an integrated transmitter. There are “plus” versions of Libre 2 and 3, both of which work for 15 days. FreeStyle Libre 2 Plus can create an AHCL system by connecting to the Omnipod 5 or Tandem t:slim X2. Libre 3 Plus makes an AHCL System Powered by Tidepool [63]. Abbott introduced its latest sensor, Instinct, in September 2025. Its dimensions are comparable to those of the Libre 3 sensor, and, like the latter, it operates for 15 days. The Instinct sensor connects to the Medtronic 780G pump to create AHCL. A software update will be required to connect the 780G pump to the Instinct sensor. Currently, the sensor is only available for sale in the U.S. [64].

Accu-Chek SmartGuide

This recently launched Roche sensor works for 14 days and comes with a single-useserter. Results are displayed in the mobile application [65]. Interestingly, the manufacturer has also invented another application, called ACCU-CHEK SmartGuide Predict. It has a function that predicts hypoglycemia within 30 min and potential changes in glucose concentration within 2 h [66]. Results obtained by the sensor can also be displayed on the Apple Watch after installing a dedicated application [67].

Medtrum

S9 CGM Sensor TouchCare is a fairly new product on the Polish market. This 14-day sensor requires a transmitter to become active. Interestingly, the device works for up to a year without needing to be charged. The manufacturer supplies the sensor and a single-useserter. The tiny size of the system is a great advantage. The manufacturer points out that the transmitter weighs only 1 g. The user can display results in a mobile application [68]. The patient’s caregivers can use the follow-up application [69]. Dedicated portals have been developed to fully track results, generate reports, and share them with medical staff [70].

Sibionics GS1

This is also an innovation on the Polish market. The sensor, which operates for 14 days, is integrated with a transmitter and inserted using a single-useserter. Results are sent to a mobile application installed on a smartphone [71]. The man-

ufacturer informs users on its website that AGP reports will be generated via the application [72].

Sinocare iCan I3

This 15-day sensor has recently been launched on the market. The sensor is integrated with a transmitter and is inserted with a single-useserter. It connects to a mobile application on a smartphone [73]. Results can be shared with therapy partners [74]. According to the manufacturer’s instructions, this sensor can be installed only on the abdominal skin.

Bioton CareSens Air

This 15-day sensor has been on the Polish market for a relatively short time. It has an integrated transmitter that communicates with the mobile application [75]. A dedicated application is designed for analytical purposes of results and for carers [76]. The website platform is also available for medical staff, carers, and users to analyze results [77].

Synoptis LynX

The Lynx sensor is a recent innovation on the Polish market. It is a 15-day sensor with a built-in transmitter. The mobile application has been designed for the user [78]. Importantly, the sensor is manufactured by MicroTech Medical, the company that also produces Equil, a drainless insulin pump available on the Polish market. Currently, there is no information on the Polish manufacturer’s website; however, on foreign portals, this pump is listed as featuring a Lynx sensor in a closed-loop system [79].

Eversense E3

It is the only subcutaneously implanted sensor that works for up to 180 days. Unlike all other mentioned sensors, its operation is not based on an enzymatic reaction, but on a fluorescence method. Implantation of the sensor is performed by medical personnel after an incision is made in the patient’s skin on the arm. After the sensor’s lifespan is complete, medical personnel also remove the device. In Poland, the first sensor was implanted in Poznań in 2017 [80]. Subsequently, after the 18.3 × 3.5 mm sensor is inserted subcutaneously, strip patches are applied to the edges of the skin incision. The sensor transmits data to a transmitter taped to the skin over it. The transmitter sends results to the mobile application, inductively charges the sensor, and is equipped with a vibration alarm, which is useful when the user does not have a smartphone [81, 82]. A dedicated application is designed for caregivers [83]. Glu-

cose results on this system can also be displayed on an Apple Watch.

Studies evaluating CGM

The effectiveness of the rtCGM system in patients using multiple daily injections (MDI) was demonstrated in the DIAMOND study [84]. This was a randomized trial of 158 patients with type 1 diabetes, using MDI, with HbA1c between 7.5% and 9.9% (59–84 mmol/mol). The subjects were divided into two groups: those using rtCGM (Dexcom G4 Platinum) and the control group, in a 2 : 1 ratio. The primary endpoint was identified with the effect on HbA1c after 24 weeks. The study showed a statistically significant decrease in HbA1c after 24 weeks, averaging 1.0% (10.9 mmol/mol) in the rtCGM group and 0.4% (4 mmol/mol) in the control group. In addition, it revealed increased TIR after 24 weeks in the CGM group, on average by 77 min/day, and reduced glycemic variability in the rtCGM group from 42% to 38%; the mean hypoglycemic time in the rtCGM group was statistically significantly shorter in comparison to the control group, i.e., 43 min/24 h vs 80 min/24 h.

The COMISAIR study demonstrated the efficacy of rtCGM compared to SMBG, regardless of the method of insulin administration [85]. Sixty-five patients with type 1 diabetes were subject to a 1-year follow-up. Patients were divided into four groups: those using rtCGM in combination with CSII, those using CGM in combination with MDI, those using SMBG in combination with CSII, and those using SMBG in combination with MDI. The primary endpoint of the study was a decrease in HbA1c. The secondary endpoint was %CV, TIR, and incidents of hypoglycemia. The groups using rtCGM (Dexcom G4) demonstrated a statistically significant decrease in HbA1c, from 8.3% (67 mmol/mol) to 7.1% (54 mmol/mol), comparable for both treatment methods. The group using CSII also exhibited improved metabolic control, i.e., from 8.4% (67 mmol/mol) to 7.9% (63 mmol/mol). TIR significantly increased in the group using rtCGM (from 50% to 69%) and in the group using CSII with SMBG (from 51% to 59%). Glycemic variability also significantly decreased in the rtCGM-using group. The rtCGM-using group also demonstrated a shorter hypoglycemia period, with a decline from 8% to 6%.

It is also essential to start using rtCGM as early as possible in new diabetes therapies. It has been confirmed that the earliest possible implementation of rtCGM systems brings the most significant benefits [86]. Three hundred and ninety-six subjects with newly diagnosed type 1 diabetes were divided into three groups: a group where rtCGM was used up to 1 year following the diagnosis, a group that did not use rtCGM, and a group

where rtCGM was implanted 3 years following the diagnosis. The patients were monitored for a period of 7 years. The observation revealed that those who used rtCGM immediately after diagnosis demonstrated a statistically significant decrease in HbA1c levels compared to those not using rtCGM. In year 7 of the observation, the difference was 7.6 (59 mmol/mol) vs. 9.8 (84 mmol/mol).

Benefits of CGM are also found in pregnant women diagnosed with diabetes. The randomized CONCEPTT study was conducted on 325 pregnant women or those who were planning pregnancy. Patients were divided into groups using rtCGM (Medtronic Guardian RT, MiniMed Minilink) and the control group not using rtCGM [87]. The researchers observed a slight difference in HbA1c levels between the rtCGM-using group, with a mean difference of 0.19% (2 mmol/mol), and the control group. The TIR in pregnant women and the control group was 68% and 61%, respectively, whereas the TAR in these groups was 27% and 32%, respectively. The number of severe hypoglycemic episodes was comparable in the CGM group and the control group (18 vs. 21). TBR was also similar, i.e. 3% vs. 4%. Neonatal health outcomes improved significantly: the number of hospitalizations in an intensive care unit lasting longer than 24 h was lower, as was the number of hypoglycemic episodes in neonates; additionally, the length of neonatal hospitalizations after birth was shorter. This study did not reveal any benefits for women using rtCGM while planning pregnancy.

It was found that the rtCGM system (Dexcom G6, G6Pro, Freestyle Libre Pro) can also be used in patients with cystic fibrosis. Screening tests for cystic fibrosis-related diabetes (CFRD) in patients with cystic fibrosis should be performed annually starting at the age of 10 years. The rtCGM system alone is suitable for metabolic control. However, CFRD cannot be diagnosed based on results obtained from CGM. The authors of a prospective study involving 77 patients with cystic fibrosis investigated the relationship between the mean HbA1c, mean rtCGM glycemia, and the diagnosis of CFRD, as well as normal or abnormal glucose tolerance [88]. Thirty-one subjects were diagnosed with CFRD; the mean rtCGM glycemia was associated with the mean HbA1c. There was greater sensitivity and specificity for rtCGM in the diagnosis of CFRD compared to HbA1c.

The possibility of diagnosing CFRD based on rtCGM (Dexcom G6, G6 Pro, Freestyle Libre Pro) was also considered, compared with oral glucose tolerance test (OGTT). In a meta-analysis, data from 1,277 publications across 19 studies were considered, involving a total of 416 patients who used rtCGM and underwent OGTT at the same time [89]. Based on the analysis, the authors concluded

that single rtCGM readings above 200 mg/dl (11.1 mmol/l) are insufficient to make an accurate diagnosis of CFRD, and further prospective studies are needed to establish cut-off points for the diagnosis of CFRD based on rtCGM data.

The safety of hospital use of rtCGM systems (Dexcom G6) was also tested. A randomized trial included 185 patients with type 1 or type 2 diabetes hospitalized in internal medicine or surgical wards [90]. Study subjects were divided into a group using an rtCGM system ($n = 83$) and the control group using SMBG and a blinded rtCGM system ($n = 79$). Both groups were treated with insulin in a basal-bolus regimen, with doses adjusted based on rtCGM or SMBG results. A significant reduction in recurrent hypoglycemia was observed in the rtCGM group (1.80 ± 1.54 vs. 2.94 ± 2.76 events per patient). Additionally, the rtCGM group, rather than the control group, demonstrated lower TBR ($1.1 \pm 3.27\%$ vs. $5.47 \pm 8.49\%$) and a lower rate of hypoglycemic episodes.

Future perspectives

Artificial intelligence opens a new era in diabetology, starting with predictive machine-learning models for risk assessment of type 2 diabetes development in individuals, based on rtCGM reports and other variables [91]. This enables early preventive intervention, finally leading to technologies, including rtCGM and AHCL systems, increasing not only the safety of type 1 diabetes management by avoiding severe hypoglycemic events, as well as decreased glucose variation and hyperglycemia or long-term complications, but also more tailored therapeutic decisions based on individual metabolic responses or backward analysis of unpleasant symptoms and acute complications with healthcare professionals [92]. According to a recent report, based on 14-day rtCGM recordings and machine learning techniques, five novel clusters of diabetes were identified, depending on the glucose pattern, in which successful treatment responses were noted for various drug classes. This sheds light on a possibly new diabetes typology [93]. Interestingly, in the last year, the rtCGM system (Dexcom Stelo) has been approved by the FDA as an over-the-counter device for individuals without diagnosed diabetes seeking information on how diet and exercise impact their glucose metabolism. Integration of Dexcom Stelo data with AI-enabled wearable technologies, such as a ring that synchronizes CGM readings with biometric and lifestyle data and formulates diet advice, illustrates a clear trend toward metabolic health monitoring and personalized health insights. However, clinical data proving benefits in this group are still limited, and this recommendation should be considered with caution [94]. The future challenge lies

in integrating clinical and rtCGM data to achieve more individualized patient analysis, where artificial intelligence (AI) and blockchain technology may provide personal security. Additionally, noninvasive rtCGM solutions, such as disposable skin patches and more user-friendly devices, are promising [95]. It is also worth paying attention to the developing needle-free injection technology (NFIT), which, in addition to reducing the risk of skin punctures, enables faster and more precise drug administration and is also suitable for patients with needle phobia [96].

Discussion

Several articles on currently available or previous versions of CGM systems have been analyzed above. The latest versions of rtCGM sensors from manufacturers such as Medtronic, Dexcom, Abbott, and the long-term sensor from Eversense are often analyzed, and their effectiveness is consistently confirmed [97–99]. There are far fewer test results for sensors being introduced to the market. Due to the emergence of new sensors, an international group of diabetes experts has issued recommendations on parameters that new sensors approved for sale in Europe should meet [4]. Unfortunately, at present, it is often difficult to find studies evaluating a particular sensor, or only the results of studies on other sensor versions are available. At the end of 2025, the PubMed database had no articles on rtCGMs, including Lynx, or S9 CGM Sensor TouchCare and only one result for iCan I3 or Sibionics GS1 (Figure 4). The websites of new manufacturers often include much less information than those of more established providers. They do not cite articles describing the sensor being sold, which the leading manufacturers usually do. Of the new sensors mentioned in this article, most information is available for the Accu-Chek SmartGuide sensor and the CareSens Air [100–103].

Before being placed on the market in European countries, medical equipment must obtain a CE certificate. The CE mark signifies that the product complies with applicable law [104]. For CGM sensors classified as IIb (moderate to high risk) products, an assessment by a notified body (a government-appointed, independent, and specialist organization that can assess a medical device) is also required [105]. Despite the need for device evaluation by a specialized organization, there are no standards for evaluating CGM sensors, nor guidelines for conducting trials correctly. The need to introduce ISO standards for CGM sensors was highlighted by the International Federation of Clinical Chemistry and Laboratory Medicine, which established an expert group to develop them [106]. During certification, a noti-




Figure 4. Distribution of PubMed-indexed publications mentioning selected continuous glucose monitoring (CGM) systems as of November 2025. The FreeStyle Libre series dominates the scientific literature, followed by Dexcom G6 and Medtronic Guardian. Other sensors, including Dexcom G7 and Eversense E3, account for a smaller proportion of indexed studies, reflecting their more recent introduction to the market

fied body may issue a CE mark based on the assessment of only a representative sample of the equipment data or when the manufacturer registers the device as “equivalent” to a device already bearing the CE mark [107]. As a result, clinical studies assessing safety can only be presented for the first generation of the device [108]. Importantly, the lack of standards defining the trial protocol and the equipment manufacturer’s sponsorship of the research may affect the reliability of the results [109]. The new EU Medical Device Regulation (MDR) 2017/745/EU, in force since 2021, was supposed to improve the certification process [110]. This document provides for greater oversight of notified bodies by member states and was intended to improve product safety and

enhance post-market product assessment. Unlike European certification, FDA registration of a medical product is more rigorous. Due to the popularization of CGM systems, a notification pathway known as the iCGM criteria has been developed. These guidelines are the first standards for study design and define measurement accuracy. The certification process in the US is carried out by a government organization, which may extend the acceptance time, but limits the risk of abuse. The process of obtaining the CE mark is presented in Table IV. Figure 5 shows the CE marking timeline for CGM systems.

Additionally, when selecting a sensor, the user should consider its functionality. Some systems are equipped with simple alarms that are suffi-

Table IV. The process of obtaining the CE mark

Stage	Place	Operation	Standards
Before certification	Manufacturer	The producer has to complete documentation of the process of designing and producing the product, clinical data, compliance of descriptive materials and labels, and potential risks of using the product, with the possibility of reporting them. Applies for CE marking to Notified body.	ISO 13485 – refers to quality management for medical devices QMS – requires clinical evaluation
Starting process	Notified body	Notified body making risk classification (Class I – low risk, IIa – medium risk, IIb – medium to high or III – high risk). Notified body conducts an audit to confirm compliance with applicable regulations. CE marking granted for 5 years, no detailed evidence published.	ISO 13485 QMS EU MDR 2017/745
CE mark received	MHRA	When a sensor obtains CE mark, the producer notifies the MHRA and presents plans for controlling the product after placing it on the market.	EU MDR 2017/745

CE – *Conformité Européenne*, EU – *European Union*, ISO – *International Organization for Standardization*, MDR – *Medical Device Regulations*, MHRA – *Medicines and Healthcare products Regulatory Agency*, QMS – *Quality Management System*.




Figure 5. MARD values along the CE-marking timeline of CGM systems. Devices introduced in 2021–2023 exhibited higher MARDs, whereas the most recent releases show decreasing MARDs, indicating improved accuracy

cient for less demanding patients, e.g., those with low glycemic variability or those treated with drugs that do not cause hypoglycemia. Patients who require intensive insulin therapy should use more sophisticated alarms. Regarding Dexcom products, the G7 sensor has significantly more alarms than the One+. Similarly, the new Accu-Chek SmartGuide sensor features a predictive application that is well suited for patients with unstable blood glucose levels [101, 102]. Hypoblocking is present in closed-loop insulin pumps (780G, Ypso) [32, 48, 49]. The previously mentioned SMART CGM systems, combined with an InPen injector, also have a prediction function [30, 39, 41]. Unfortunately, the InPen injector is currently an expensive solution, so few patients decide to use it. When choosing the most suitable system with available alarms, the user should consider the possibility of “alarm fatigue” [111]. In a survey of individuals who stopped using the rtCGM system, 36% reported too-frequent alarms and 46% reported false alarms [112]. The most commonly cited reasons are accuracy, followed by inconvenience, price, and lack of insurance coverage. Therefore, it is essential to inform the patient about the lag-time phenomenon, which results from differences in measurement locations [9, 11].

Physically active patients may require rtCGM systems equipped with features useful during sports. Using an rtCGM sensor may be challenging when the patient needs to carry a phone or an insulin pump at all times to obtain results. The mentioned Dexcom systems will work well in this solution, as they can receive data directly on

a smartwatch [55, 56]. The Eversense system is an excellent solution for patients training in contact combat sports or those exposed to water for extended periods. They will not have to worry about the sensor being accidentally removed, as it is implanted subcutaneously [82]. Some patients experience sensor detachment. Hence, some manufacturers provide dedicated patches to protect the sensor. Sensor supporting bands or adhesive bandages can also be helpful.

Conclusions

The emergence of rtCGM sensors on the market provides healthcare providers with insightful tracking of patients' disease course, glucose self-management, and lifestyle, enabling a personalized approach and tailored treatment. Physicians should respond to these adjustments by expanding and updating their knowledge of available rtCGM devices, their cost-effectiveness, and the results of clinical studies. Available sensors exhibit very similar MARD values. Therefore, when choosing an rtCGM device, we should focus more on the features discussed in the article than on a nonsignificant difference in measurement error. Manufacturers of new rtCGM systems should be obligated by policymakers to conduct extensive peer-reviewed research and analysis to help medical professionals become more familiar with the characteristics of specific sensors. There is also an increasing need for independent, peer-reviewed evaluations of new rtCGM parameters to form authoritative recommendations from diabetes associations. Ensuring equitable access to CGM requires transparent reimbursement criteria, tar-

SWOT: Adoption of CGM (2025)




Figure 6. SWOT Analysis of CGM adoption (2025)

AHCL – advanced hybrid closed loop, CGM – continuous glucose monitoring, HCPs – healthcare professionals, MDI – multiple daily injections, rtCGM – real-time CGM.

geted support for underserved groups, and integration of digital literacy training into routine care. Additionally, new solutions and AI implementations enable manufacturers of competing systems to upgrade their technology and offer new solutions continually. Currently, only three hybrid insulin pumps are officially available in Poland; however, this article reports that new AHCL solutions are likely to become available soon. A SWOT analysis of rtCGM adoption is provided in Figure 6.

Funding

No external funding.

Ethical approval

Not applicable.

Conflict of interest

The authors declare no conflict of interest.

References

1. Bender C, Vestergaard P, Cichosz SL. The history, evolution and future of continuous glucose monitoring (CGM). *Diabetology* 2025; 6: 17.
2. Abbott. Real-world data show Abbott's FreeStyle Libre® systems and GLP-1 medicines work better together for people with type 2 diabetes. Abbott MediaRoom. 2024 Mar 6. Available from: <https://www.abbott.co.uk/media-center/news/real-world-data-show-abbotts-free-style-libre-systems-and-glp-1-medicines-work-better-together-for-people-with-type-2.html>

3. Regulation of the Minister of Health of 13 October 2023 amending the regulation on the list of medical devices issued on commission. Available from: <https://isap.sejm.gov.pl/isap.nsf/download.xsp/WDU20230002461/O/D20232461.pdf> Accessed August 05, 2025.
4. Mathieu C, Itrace C, Wilmot EG, et al. Minimum expectations for market authorization of continuous glucose monitoring devices in Europe –'eCGM' compliance status. *Diabetes Obes Metab* 2025; 27: 1025-31.
5. Mysliwiec M, Czupryniak L, Gellert R, et al. Continuous glucose monitoring systems can meet the challenge of glucose management and beyond in individuals with type 2 diabetes: an expert multidisciplinary position. *Diabetes Therapy* 2025; 16: 1745-61.
6. Glukometr.eu – comparison of glucometer specifications, prices and news. Available from: <https://glukometr.eu/porownywarka-glukometrow/> Accessed August 05, 2025.
7. International Organization for Standardization. In vitro diagnostic test systems – requirements for blood-glucose monitoring systems for self-testing in managing diabetes mellitus. EN ISO 15197:2015. Geneva, 2015.
8. Basu A, Dube S, Slama M, et al. Time lag of glucose from intravascular to interstitial compartment in humans. *Diabetes* 2013; 62: 4083-7.
9. Schmelzeisen-Redeker G, Schoemaker M, Kirchsteiger H, Freckmann G, Heinemann L, Re LD. Time delay of CGM sensors: relevance, causes, and countermeasures. *J Diabetes Sci Technol* 2015; 9: 1006-15.
10. Rodbard D. Continuous glucose monitoring: a review of successes, challenges, and opportunities. *Diabetes Technol Ther* 2016; 18: S3-13.

11. Basu A, Dube S, Veetil S, et al. Time lag of glucose from intravascular to interstitial compartment in type 1 diabetes. *J Diabetes Sci Technol* 2015; 9: 63-8.
12. Kovatchev BP, Patek SD, Ortiz EA, Breton MD. Assessing sensor accuracy for non-adjunct use of continuous glucose monitoring. *Diabetes Technol Ther* 2015; 17: 177-86.
13. Battelino T, Danne T, Bergenstal RM, et al. Clinical targets for continuous glucose monitoring data interpretation: recommendations from the international consensus on time in range. *Diabetes Care* 2019; 42: 1593-603.
14. Johnson ML, Martens WT, Criego AB, Carlson AL, Simonson GD, Bergenstal RM. Utilizing the ambulatory glucose profile to standardize and implement continuous glucose monitoring in clinical practice. *Diabetes Technol Ther* 2019; 21: 217-25.
15. Danne T, Nimri R, Battelino T, et al. International consensus on use of continuous glucose monitoring. *Diabetes Care* 2017; 40: 1631-40.
16. Battelino T, Alexander CM, Amiel SA, et al. Continuous glucose monitoring and metrics for clinical trials: an international consensus statement. *Lancet Diabetes Endocrinol* 2023; 11: 42-57.
17. Monnier L, Colette C, Wojtusciszyn A, et al. toward defining the threshold between low and high glucose variability in diabetes. *Diabetes Care* 2017; 40: 832-8.
18. Louis Monnier L, Colette C, Owens D. The glycemic triumvirate and diabetic complications: is the whole greater than the sum of its component parts? *Diabetes Res Clin Pract* 2012; 95: 303-11.
19. Bergenstal RM, Beck RW, Close KL, et al. Glucose Management Indicator (GMI): a new term for estimating A1c from continuous glucose monitoring. *Diabetes Care* 2018; 41: 2275-80.
20. American Diabetes Association Professional Practice Committee. 2. Diagnosis and classification of diabetes: Standards of Care in Diabetes-2025. *Diabetes Care* 2025; 48 (Suppl 1): S27-49.
21. Zhang Z, Wang Y, Lu J, Zhou J. Time in tight range: a key metric for optimal glucose control in the era of advanced diabetes technologies and therapeutics. *Diabetes Obes Metab* 2025; 27: 450-6.
22. Passanisi S, Piona C, Salzano G, et al. Aiming for the best glycemic control beyond time in range: time in tight range as a new CGM metric in children and adolescents with type 1 diabetes using different treatment modalities. *Diabetes Technol Ther* 2023; 26: 161-6.
23. Dunn TC, Ajjan RA, Bergenstal RM, Xu Y. Is It time to move beyond TIR to TITR? Real-world data from over 20,000 users of continuous glucose monitoring in patients with type 1 and type 2 diabetes. *Diabetes Technol Ther* 2024; 1: 203-10.
24. Czupryniak L, Dzida G, Fichna F, et al. Ambulatory glucose profile (AGP) report in daily care of patients with diabetes: practical tips and recommendations. *Diabetes Ther* 2022; 13: 811-21.
25. Clinical recommendations of the Polish Diabetes Association. Defining goals in diabetes treatment. *Curr Top Diabetes* 2025; 5: 21-4.
26. Clinical recommendations of the Polish Diabetes Association. General principles of treatment for people with type 1 diabetes. *Curr Top Diabetes* 2025; 5: 19-21.
27. Clinical recommendations of the Polish Diabetes Association. Glycemic monitoring. *Curr Top Diabetes* 2025; 5: 42-7.
28. Clinical Recommendations of the Polish Diabetes Association. Diabetes in Children and Adolescents. *Curr Top Diabetes* 2025; 5: 103-7.
29. Medtronic Diabetes Poland. System SMART CGM Guardian 4. Available from: <https://www.medtronic-diabetes.com/pl-PL/guardian4-cgm-system>. Accessed August 05, 2025.
30. Medtronic Diabetes Poland. System MDI, Simplera. Available from: <https://www.medtronic-diabetes.com/pl-PL/SmartMDI>. Accessed August 05, 2025.
31. Medtronic Diabetes. User guide Guardian 4. Available from: <https://www.medtronicdiabetes.com/download-library/guardian-4-sensor-transmitter>
32. MedtronicDiabetesPoland.System780G.Availablefrom: <https://www.medtronic-diabetes.com/pl-PL/terapia-z-uzyciem-pompy-insulinowej/pompa-insulinowa-minimed-780g>. Accessed August 05, 2025.
33. MiniMed™ Mobile. App Google Play https://play.google.com/store/apps/details?id=com.medtronic.diabetes.minimedmobile.eu&pcampaignid=web_share Accessed August 05, 2025
34. MiniMed™ Mobile. App Store <https://apps.apple.com/pl/app/minimed-mobile/id1464664725?l=pl> Accessed August 05, 2025.
35. CareLink™ Connect. App Google Play. <https://play.google.com/store/apps/details?id=com.medtronic.diabetes.carepartner&hl=pl> Accessed August 05, 2025.
36. CareLink™ Connect. App Store <https://apps.apple.com/pl/app/carelink-connect/id1467892577?l=pl> Accessed August 05, 2025.
37. CareLink™ Personal. Medtronic Diabetes Poland. <https://www.medtronic-diabetes.com/pl-PL/wsparcie/carelink> Accessed August 05, 2025.
38. CareLink™ System – For healthcare professionals only. <https://carelink.medtronic.eu/login> Accessed August 05, 2025.
39. Simplera™ App Google Play. https://play.google.com/store/apps/details?id=com.medtronic.diabetes.simplera.eu&pcampaignid=web_share Accessed August 05, 2025.
40. Simplera™ App Store. <https://apps.apple.com/pl/app/simplera/id1666540280?l=pl> Accessed August 05, 2025.
41. InPen, Diabetyk24 online shop. <https://diabetyk24.pl/inteligentny-wstrzykiwacz-insuliny-inpentm-1-szt-mmt-105nnblw1> Accessed August 05, 2025.
42. InPen App Google Play. https://play.google.com/store/apps/details?id=com.companionmedical.inpen&pcampaignid=web_share Accessed August 05, 2025.
43. InPen App Store. <https://apps.apple.com/pl/app/intelligentny-pen-ins-inpen/id1573954312?l=pl> Accessed August 05, 2025.
44. Silva JD, Lepore G, Battelino T, et al. Real-world performance of the MiniMed™ 780G system: first report of outcomes from 4120 users. *Diabetes Technol Ther* 2022; 24: 113-9.
45. Dexcom G6 App Google Play. https://play.google.com/store/apps/details?id=com.dexcom.g6.region2.mgdl&pcampaignid=web_share Accessed August 05, 2025.
46. Dexcom Follow App Google Play. https://play.google.com/store/apps/details?id=com.dexcom.follow.region2.mgdl&pcampaignid=web_share Accessed August 05, 2025.
47. Dexcom Clarity App Google Play. https://play.google.com/store/apps/details?id=com.dexcom.clarity.mobile&pcampaignid=web_share Accessed August 05, 2025.

48. YpsoPump with mylife Loop – mylife Poland. <https://www.mylife-diabetescare.com/pl-PL/system-mylife-loop.html> Accessed August 05, 2025.
49. mylife CamAPS FX App Google Play. https://play.google.com/store/apps/details?id=com.camdiab.fx.alert.mgdl&pcampaignid=web_share Accessed August 05, 2025.
50. Alwan H, Wilinska ME, Ruan Y, Silva JD, Hovorka R. Real-world evidence analysis of a hybrid closed-loop system. *J Diabetes Sci Technol* 2025; 19: 385-9.
51. Dexcom G7 App Google Play. https://play.google.com/store/apps/details?id=com.dexcom.g7&pcampaignid=web_share Accessed August 05, 2025.
52. Dexcom One+ App Google Play. https://play.google.com/store/apps/details?id=com.dexcom.d1plus&pcampaignid=web_share Accessed August 05, 2025.
53. DexcomOne+ receiver, Diabetyk24 Online shop. <https://diabetyk24.pl/odbiornik-dexcom-one> Accessed August 05, 2025.
54. Website Dexcom. News. <https://www.dexcom.com/g7-15-day-sensor-fda> Accessed August 05, 2025.
55. Website diaTribe. <https://diatribe.org/diabetes-technology/dexcom-launches-direct-watch-g7-now-connects-directly-apple-watch-us> Accessed August 05, 2025.
56. Website Garmin. News. <https://www.garmin.com/en-US/newsroom/press-release/wearables-health/people-with-diabetes-can-now-view-dexcom-cgm-data-on-their-garmin-smartwatch-or-cycling-computer/> Accessed August 05, 2025.
57. FreeStyle LibreLink App Google Play. https://play.google.com/store/apps/details?id=com.freestylelibre.app.pl&pcampaignid=web_share Accessed August 05, 2025.
58. Website Abbott. Online shop. <https://www.freestyle.abbott/pl-pl/produkty/sensor-freestyle-libre-2.html> Accessed August 05, 2025.
59. LibreLinkUp App Google Play. https://play.google.com/store/apps/details?id=org.nativescript.LibresLinkUp&pcampaignid=web_share Accessed August 05, 2025.
60. Website LibreView. <https://www.libreview.com/> Accessed August 05, 2025.
61. Website Abbott. FreeStyle Libre 3. <https://www.freestyle.abbott/us-en/products/freestyle-libre-3.html> Accessed August 05, 2025.
62. Website Abbott. Automated insulin delivery system with Libre 3. <https://www.freestyle.abbott/uk-en/products/connected-care/mylife-loop-aid-partnership.html> Accessed August 05, 2025.
63. Website Abbott. Partner Integrations. <https://www.diabetescare.abbott/support/partnerships.html> Accessed September 30, 2025.
64. Website Medtronic. Introducing the new Instinct sensor, made by Abbott. <https://www.medtronicdiabetes.com/loop-blog/introducing-instinct-sensor> Accessed September 30, 2025.
65. SmartGuide – Accu-Chek CGM App Google Play. https://play.google.com/store/apps/details?id=com.roche.smartguide.eu&pcampaignid=web_share Accessed August 05, 2025.
66. Predict – ACCU-CHEK SmartGuide App Google Play. https://play.google.com/store/apps/details?id=com.roche.smartguide.predict.eu&pcampaignid=web_share Accessed August 05, 2025.
67. ACCU-CHEK® SmartGuide App Store. <https://apps.apple.com/de/app/accu-chek-smartguide-app/id6503455257?l=en-GB> Accessed August 05, 2025.
68. Medtrum EasySense App Google Play. https://play.google.com/store/apps/details?id=com.medtrum.easysenseforandroidmg&pcampaignid=web_share Accessed August 05, 2025.
69. Medtrum EasyFollow App Google Play. https://play.google.com/store/apps/details?id=com.medtrum.easyfollowforandroidmg&pcampaignid=web_share Accessed August 05, 2025.
70. Website TouchCareCGM. Systems: EasyView Personal and Pro. <https://touchcarecgm.pl/system-easyview/> Accessed August 05, 2025.
71. SIBIONICS App Google Play. https://play.google.com/store/apps/details?id=com.sisensing.sijoy&pcampaignid=web_share Accessed August 05, 2025.
72. Website SibioniscGm. <https://www.sibioniscgcm.com/pl-ro/products/sibionics-gs1-continuous-glucose-monitoring-cgm-system-2> Accessed August 05, 2025.
73. iCan CGM App Google Play. https://play.google.com/store/apps/details?id=com.sinocare.cgm.ce&pcampaignid=web_share Accessed August 05, 2025.
74. iCan Reach App Google Play. https://play.google.com/store/apps/details?id=com.sinocare.cgm.follow.ce&pcampaignid=web_share Accessed August 05, 2025.
75. CareSens Air App Google Play. https://play.google.com/store/apps/details?id=com.isens.csair&pcampaignid=web_share Accessed August 05, 2025.
76. Sens365 App Google Play. https://play.google.com/store/apps/details?id=com.isens.sens365&pcampaignid=web_share Accessed August 05, 2025.
77. Website Sens365. I-Sens. <https://sens365.i-sens.com/home> Accessed August 05, 2025.
78. LinX CGM App Google Play. https://play.google.com/store/apps/details?id=com.microtech.aidexx.mgdl&pcampaignid=web_share
79. Website Lynxcgms.com LynxGlobal. <https://lynxcgms.com/products/equil-patch-insulin-pump?srsltid=AfmBOoryR0VER3SIY-GRSEMYthVoxqOqGbkY9g8Dzq8FNzGeb5j43XZp> Accessed August 05, 2025.
80. Website mojacukrzyca.org, Interview with the first patient in Poland with an implanted glucose monitor. <https://www.mojacukrzyca.org/?a=text&id=4468&des=wywiad-z-pierwsza-pacjentka-w-polsce-z-wszczepionym-monitoringiem-glikemii> Accessed August 05, 2025.
81. Eversense CGM App Google Play. https://play.google.com/store/apps/details?id=com.senseonics.gen12androidapp&pcampaignid=web_share Accessed August 05, 2025.
82. Website Ascensia Diabetes Care Poland. <https://www.diabetes.ascensia.pl/eversense/eversense-e3/> Accessed August 05, 2025.
83. Eversense NOW App Google Play. https://play.google.com/store/apps/details?id=com.senseonics.careapp&pcampaignid=web_share Accessed August 05, 2025.
84. Beck RW, Riddlesworth T, Ruedy K, et al. Effect of continuous glucose monitoring on glycemic control in adults with type 1 diabetes using insulin injections: the DIAMOND randomized clinical trial. *JAMA* 2017; 317: 371-8.
85. Šoupal J, Petruželková L, Flekač M, et al. Comparison of different treatment modalities for type 1 diabetes, including sensor-augmented insulin regimens, in 52 weeks of follow-up: a COMISAIR study. *Diabetes Technol Ther* 2016; 18: 532-8.
86. Champakanath A, Akturk HK, Alonso GT, Snell-Bergeon JK, Shah VN. Continuous glucose monitoring initiation

- within first year of type 1 diabetes diagnosis is associated with improved glycemic outcomes: 7-year follow-up study. *Diabetes Care* 2022; 45: 750-3.
87. Feig DS, Donovan LE, Corcoy R, et al. Continuous glucose monitoring in pregnant women with type 1 diabetes (CONCEPTT): a multicentre international randomised controlled trial. *Lancet* 2017; 390: 2347-59.
 88. Scully KJ, Sherwood JS, Martin K, et al. Continuous glucose monitoring and HbA1c in cystic fibrosis: clinical correlations and implications for CFRD diagnosis. *J Clin Endocrinol Metab* 2022; 107: 1444-54.
 89. Kumar S, Pallin M, Soldatos G, Teede H. Comparison of continuous glucose monitoring to reference standard oral glucose tolerance test for the detection of dysglycemia in cystic fibrosis: a systematic review. *J Clin Transl Endocrinol* 2022; 27: 100305.
 90. Spanakis EK, Urrutia A, Galindo RJ, et al. Continuous glucose monitoring-guided insulin administration in hospitalized patients with diabetes: a randomized clinical trial. *Diabetes Care* 2022; 45: 2369-75.
 91. Yang X, Yao M, Huang J, Cheng Z, Sun T. Machine learning predicts diabetes risk in high-risk populations: based on the National Health and Nutrition Examination Survey database. *Arch Med Sci* 2026; 22L 708-20.
 92. Fleischer J, Hansen TK, Cichosz SL. Hypoglycemia event prediction from CGM using ensemble learning. *Front Clin Diabetes Healthc* 2022; 3: 1066744.
 93. 61st EASD Annual Meeting of the European Association for the Study of Diabetes. *Diabetologia* 2025; 68 (Suppl 1): 1-754.
 94. U.S. Food and Drug Administration. FDA clears first over-the-counter continuous glucose monitor. Press announcement. March 05, 2024. Available from: <https://www.fda.gov/news-events/press-announcements/fda-clears-first-over-counter-continuous-glucose-monitor>. Accessed September 18, 2025.
 95. Ji C, Jiang T, Liu L, Zhang J, You L. Continuous glucose monitoring combined with artificial intelligence: re-defining the pathway for prediabetes management. *Front Endocrinol* 2025; 16: 1571362.
 96. Bafna HR, Shirsath NR. Needle-free injection technology as a new approaches for drug delivery. *Med Pharm J* 2025; 4: 134-54.
 97. Hohendorff J, Czupryniak L, Dzida G, et al. Novel methods of continuous glucose monitoring and telehealth in the improvement of diabetes care: a narrative review. *Arch Med Sci* 2023; 19: 757-64.
 98. Freckmann G, Wehrstedt S, Eichenlaub M, et al. A comparative analysis of glycemic metrics derived from three continuous glucose monitoring systems. *Diabetes Care* 2025; 48: 1213-7.
 99. Rohner DG, Burget L, Henzen C, Fischli S. Impact on diabetes control and patient-reported outcomes of a newer implantable continuous glucose monitoring system (Eversense® CGM System): a single-centre retro- and prospective observational study. *Swiss Med Wkly* 2024; 20: 3366.
 100. Kim KS, Lee SH, Yoo WS, Park CY. Accuracy and safety of the 15-day CareSens Air continuous glucose monitoring system. *Diabetes Technol Ther* 2024; 26: 222-8.
 101. Jendrike N, Eichenlaub M, Link M, et al. Comparative performance analysis of manual and updated optional calibration algorithms for the CareSens Air CGM system. *J Diabetes Sci Technol* 2025; 19: 1-8.
 102. Kulzer B, Heinemann L. Predicting glucose values: a new era for continuous glucose monitoring. *J Diabetes Sci Technol* 2024; 18: 1000-3.
 103. Glatzer T, Ringemann Ch, Militz D, Mueller-Hoffmann W. Concept and implementation of a novel continuous glucose monitoring solution with glucose predictions on board. *J Diabetes Sci Technol* 2024; 18: 1004-8.
 104. European Commission. CE marking. https://single-market-economy.ec.europa.eu/single-market/ce-marking_en Accessed November 07, 2025.
 105. Pemberton J, Wilmot E, Barnard-Kelly K, et al. CGM accuracy: Contrasting CE marking with the governmental controls of the USA (FDA) and Australia (TGA): a narrative review *Diabetes Obes Metab* 2023; 25: 916-39.
 106. International Federation of Clinical Chemistry and Laboratory Medicine . Working Group on Continuous Glucose Monitoring (WG-CGM) – IFCC. 2024 <https://ifcc.org/ifcc-scientific-division/sd-working-groups/wg-cgm/> Accessed November 07, 2025.
 107. Hulstaert F, Pouppez C, Primus-de Jong C, et al. Gaps in the evidence underpinning high-risk medical devices in Europe at market entry, and potential solutions. *Orphanet J Rare Dis* 2023; 18: 212.
 108. Sauerland S, Brockhaus AC, Fujita-Rohwerder N, Saad S. Approaches to assessing the benefits and harms of medical devices for application in surgery. *Langenbeck's Arch Surg* 2014; 399: 279-85.
 109. Lundh A, Lexchin J, Mintzes B, et al. Industry sponsorship and research outcome. *Cochrane Database Syst Rev* 2017; 2: MR000033.
 110. European Commission. Regulation (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices, amending Directive 2001/83/EC, Regulation (EC) No 178/2002 and Regulation (EC) No 1223/2009 and repealing Council Directives 90/385/EEC and 93/42/EEC 2017 <https://eur-lex.europa.eu/legal-content/EN/TXT/?uri=CELEX:02017R0745-20200424>. Accessed November 07, 2025.
 111. Shivers JP, Mackowiak L, Anhalt H, Zisser H. "Turn it off!": diabetes device alarm fatigue considerations for the present and the future. *J Diabetes Sci Technol* 2013; 7: 789-94.
 112. Engler R, Routh TL, Lucisano JY. Adoption barriers for continuous glucose monitoring and their potential reduction with a fully implanted system: results from patient preference surveys. *Clin Diabetes* 2018; 36: 50-8.